

Lab : 1054 / KRSNAA DIAGNOSTICS,LTD,MOGA

Cycle + Sample No : 23031004

Analyte Scheme : Haematology - CBC

PT Result Received Date : 08/12/2023 Name Of The Test : Total Leucocytes (WBC) count

Result : 3.52 Z Score : -7.30 Upload Date : 15/01/2024

**Previous trends/Unacceptable results :**

**last 6-12 month** : No Trends in last 6 months

**Last 3 months** : No Trends in last 3 months

**Data Entry Check** : Unit wrongly Assigned

**Investigation** : Raw data verify Unit was miss match .

**Conclusion** : Unit Wrong mention On portal

**Was patient data affected?** : No

**Classification Of Problem** : Technical

**Immediate Action** : Correction in unit

**Corrective Action** : From next samples maker and checker system implemented

**Preventive Action** : Check the Receive data before submission

**Remark** : After unit change the Z score is within acceptable Limits