



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Bilirubin Total	Diasys 200 Pro	0.62	mg/dl	3.94
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
# Urea	mg/dl	19.90	20.87		-0.08
@ Creatinine	mg/dl	0.73	0.83	-0.38	-0.93
X Bilirubin Total	mg/dl	0.62	0.30	3.94	1.90
@ Total Protein	g/dL	6.01	5.76	0.71	0.67
@ Albumin	g/dL	3.10	3.59	-0.49	-0.66
@ Uric Acid	mg/dl	3.36	3.74	-1.18	-1.12
@ Calcium	mg/dl	7.34	7.31	0.39	0.52
@ Cholesterol	mg/dl	122.00	130.32	-0.85	-0.65
@ Triglycerides	mg/dl	109.00	118.41	-1.54	-1.36
@ HDL	mg/dl	31.00	33.30	-0.19	-0.52
@ LDL	mg/dl	75.67	73.11	0.28	0.16
@ ALT	U/L	15.00	13.62	0.48	0.24
@ AST	U/L	24.21	26.43	-0.33	-0.46
@ ALP	U/L	43.26	42.72	-0.24	-0.04

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	75.39	81.59	-0.78	-0.67
@ Sodium	mmol/L	108.64	106.54	-0.06	-0.05
@ Potassium	mmol/L	3.40	3.48	-0.36	-0.17

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	1
Evaluated Parameters	16
Outlier Parameters (X)	1
EQAS Score Biochemistry	93.75 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070