



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 04

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.46	1.45	-0.18	0.01
@ Total Leucocytes (WBC) count	/c.mm	3360.00	3456.41	-0.20	-0.13
@ Haemoglobin	g/dL	3.90	4.05	-0.61	-0.52
@ MCV (Mean Corpuscular Volume)	fL	122.80	124.09	-0.20	-0.04
@ MCH (Mean Corpuscular Hb)	pg	26.80	27.88	-0.22	-0.42
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.80	22.18	-0.06	-0.32
@ Platelet Count	1000/ μ L	140.00	151.44	-0.15	0.02
@ Hematocrit	%	17.90	17.87	-0.23	0.06
@ RDW CV%	%	22.00	22.46	0.06	-0.96

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070