

Lab : 1050 / KRSNAA DIAGNOSTICS LTD HOSHIARPUR

Cycle + Sample No : 23031001

Analyte Scheme : **Biochemistry**

PT Result Received Date	: 08/12/2023	Name Of The Test	: Bilirubin Total
Result	: 0.62	Z Score	: 3.94
		Upload Date	: 02/01/2024

Previous trends/Unacceptable results :

last 6-12 month	: NO
Last 3 months	: NO
Data Entry Check	: CHECK CALLIBRATION
Investigation	: RERUN AND RAW DATA MAINTAIN
Conclusion	: AFTER RERUN WITH IN Z SCORE
Was patient data affected?	: NO
Classification Of Problem	: Technical
Immediate Action	: RERUN
Corrective Action	: RERUN SAMPLE
Preventive Action	: CHECK CALLIBRATION BEFORE TESTING
Remark	: Z SCORE WITH IN RANGE1.80