



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07001F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	230926.1	DT - 26 / 09 / 2023		ULR No.		CC2646230000	07001F
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	26.09.2023		
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)			Recommended Date for Next Callb.	25.09.2024		
				Certificate Issue Dt.	29.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Centrifuge M/c.	0 to 6000 RPM	10 RPM	Remi -----	KDDHPBCPCF07 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value UUC (in RPM)	Master Gauge Value (in RPM)	Deviation (in RPM)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in RPM)	
1	100	100.1	-0.1		5.88
2	500	500.2	-0.2		5.88
3	1000	1000.2	-0.2		5.88
4	3000	3000.4	-0.4		6.15
5	6000	6000.6	-0.6		6.17

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com