



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.0F-01



Certificate No	AACPL / 07010F	Field	Mech.
SRF No & Date	230926.2 DT - 26 / 09 / 2023	Comparison Method	Group
Calibration Method :		M/s Krsnaa Diagnostics Ltd.	Volume
Company Name	DH Civil Hospital, Jalandhar-Hoshiarpur	Address	Page No.
Address	Road, Hoshiarpur (PB)	ULR No.	1 of 1
		Accuracy	CC2646230000 07010F
		Calibration Date	26.09.2023
		Recommended Date for Next Calib	25.09.2024
		Certificate Issue Dt.	29.09.2023
		UUC Condition	Good

Calibration Instrument Detail			
SI	Instrument Name	Range	Least Count
1	Micro Pipette	20 to 200 µl	1 µl
		Make / SI.No.	Party ID / Location
		I-Pitte 52670	KDDHPBCCMP09 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)			
SI	Instrument Name	Make/SI No	Calibrated By
1	Weighing Balance	Radwag / 668683	Bluebox Techno
		Cal Certificate No	Cal. Valid Upto
		BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result				
S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	20	20.00052	-0.00052	0.25
2	100	100.00196	-0.00196	0.25
3	200	200.00325	-0.00325	0.25

- Note**
- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
 - 2 This certificate cannot be reproduced except in full without our prior permission in writing
 - 3 This certificate refers only to the particular item(s) submitted for calibration
 - 4 UUC - Unit under calibration

----- End of Certificate -----

Ashish
Calibrated by



Approved by
Anand Chaurasia
Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
E-mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com