



# AACPL CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.0F-01



CC - 254B

Certificate No	AACPL / 07007F	Field	Mech
SRF No & Date	230926.2	DT - 26 / 09 / 2023	
Calibration Method :	Comparison Method		
Company Name	M/s Krsnaa Diagnostics Ltd.		
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)		
Group	Volume	Page No.	1 of 1
ULR No.	CC2646230000 07007F		
Accuracy			
Calibration Date	26.09.2023		
Recommended Date for Next Callb	25.09.2024		
Certificate Issue Dt.	29.09.2023		
UUC Condition	Good		

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	10 µl	I-Pitte NA	KDDHPBCCMP10 Bio Chemistry

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

S. No	Nominal Value on UUC (in µl)	Calibration Result		
		Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in µl)
1	100	100.00159	-0.00159	0.25
2	500	500.00854	-0.00854	0.25
3	1000	1000.01293	-0.01293	0.25

#### Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

*Aswist*  
Calibrated by



Approved by  
*Apand Chaurasia*  
Apand Chaurasia  
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012  
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597  
E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com