



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07004F	Field	Thermal
SRF No & Date	230926.1	DT - 26 / 09 / 2023	
Calibration Method :	Comparison Method		
Company Name	M/s Krsnaa Diagnostics Ltd.		
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)		
Group	Temp.	Page No.	1 of 1
ULR No.	CC2646230000 07004F		
Accuracy			
Calibration Date	26.09.2023		
Recommended Date for Next Callb.	25.09.2024		
Certificate Issue Dt.	29.09.2023		
UUC Condition	Good		

SI	Instrument Name	Range	Least Count	Make/SI.No.	Party ID / Location
1	Incubator	0 to 110 °C	1 °C	Arba Healthcare	KDDHPBCCIN16 Bio Chemistry

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Digital Multimeter	Gwinstek EL 160577	AACPL	AACPL/04256F	11/07/2024
2	RTD (4 Wire) Pt-100	Tempens/ 115	AACPL	AACPL/19820F	28/09/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
DKD-R5-7	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure
 Comparison With Standard Sensors as per Calibration Procedure Manual AACPL / CP / 02, Section No : CPTH - 09
 Calibration Points :- 10 , 60 & 110 °C

I No.	Master Value		UUC Value		Deviation In °C	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (°C)
	In °C		In °C			
1	10.05		10		-0.05	0.63
2	60.11		60		-0.11	0.63
3	110.13		110		-0.13	0.63

Note
 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
 This certificate cannot be reproduced except in full without our prior permission in writing
 This certificate refers only to the particular item(s) submitted for calibration
 UUC - Unit under calibration
 Temperature Scale: International Temperature Scale Of 1990 (ITS-90)

----- End of Certificate -----



Approved by

 Anand Chaurasia
 (CEO)

Calibrated by

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
 Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
 E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com