



Calibration Certificate

Customer Name : Dr. Bharti Path Lab

Address of Customer:

Model : Erba Chem-7

Serial No. : S210821

Calibration Done Date : 24/5/2024

Calibration Due Date : On or Before 23/5/2025

This is to certify that we have calibrated the Semi Automatic Biochemistry Analyzer Model: Chem-7 bearing Sr. No. S210821 and on inspection of the instrument gains. It is observed that the results of IQC are well within the range and instrument is working fine.

Thanking

you,For

Transasia Bio-Medicals Ltd.

Devender Singh

Sr. Service Engineer

**Calibration Data Gain Values**

Filter	340	405	450	505	546	578	630	670	
Gain Values	06077	1952 6	29013	3148 8	3647 7	13729	32216	21780	
OK/ Not OK	OK	OK	OK	OK	OK	OK	OK	OK	

Temperature

RCT Temperature	37 °C	Range 36.8°C - 37.2°C
--------------------	-------	--------------------------

Voltage

SMPS voltage	17.8v
Logic Voltage	4.9v
Lamp Voltage	11.68v

**Cuvette
Blank**

Range (0.01-0.20)	All Cuvettes Absorbance's are within the Range
-------------------	---

Thanking

you,For

Transasia Bio-Medicals Ltd.

Devender Singh

Sr.Service Engineer