

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Accepted Value
X C-Reactive Protein Qualitative	mg/L	Normal	As per Reference

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	80.00 %

Checked


Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard Un
X C-Reactive Protein Qualitative	Erba EM 200	Normal	mg/L
X Widal test - Typhi O antigen (TO)	Manual	1:80	-

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Problem Classification: Widal and CRP @ Unacceptable.

Corrective Action: 1@ CRP & Widal Method needs to be change the software. Chemical. Next cycle results awaited. Training for results only done.

Reviewed by: Ashwan

Dated: 14.02.2024

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
# HBsAg	IU/mL	Reactive	
# HCV	IU/mL	Reactive	
@ HIV Ab	S/CO	Non-reactive	Non-reactive

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Accepted Value
@ C-Reactive Protein Qualitative	mg/L	Abnormal	Abnormal

Legend @ : Acceptable


Ⓜ : Acceptable

X : Unacceptable

: Not Evaluated

Ⓞ : Delayed Result Entry

Total Parameters	:
Not Evaluated Parameters	:
Evaluated Parameters	:
Outlier Parameters (X)	:
EQAS Score Immunoassay Extended	100.00 %


Seen. 
 14/02/24

Note CRP & HIV - Acceptable

for this Sample No. HBsAg & HCV
 are not evaluated. By EQAS
 Team. for this Sample No.

Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070


 14.02.24