


Instrument : ERBA H 560

Analyte	Standard Unit	Result Value	Mean	Z-Score	RI
@ Erythrocyte (RBC) Count	mill/cu.mm	2.13	2.11	0.64	0.1
* Total Leucocytes (WBC) count	/c.mm	3.73	3730.00	-11.12 *	0.2
@ Haemoglobin	g/dL	6.00	6.00	0.12	0.1
@ MCV (Mean Corpuscular Volume)	fL	142.20	136.75	1.09	0.1
@ MCH (Mean Corpuscular Hb)	pg	28.00	28.45	-0.99	-0.1
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	19.70	21.00	-1.96	-0.2
@ Platelet Count	1000/ μ L	124.00	118.00	0.66	1.1
@ Hematocrit	%	30.20	28.84	1.04	0.1
@ RDW CV%	%	21.60	20.86	0.25	-0.1

Legend @ : Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⊕ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	!
Not Evaluated Parameters	!
Evaluated Parameters	!
Outlier Parameters (X)	!
EQAS Score Haematology - CBC	88.89 %

Checked

 14/02/24

Dr Puneet Kumar Nigam


Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
X Neutrophil.	-	<40%	40-70
# Lymphocyte.	-	>40%	
X Eosinophil.	-	6-40%	<6

- Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	:
Not Evaluated Parameters	:
Evaluated Parameters	:
Outlier Parameters (X)	:
EQAS Score Haematology-PS Slide Count Qualitative	0.00 %

Checked

 14/02/24

Dr Puneet Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070