

Analyte	Instrument	Result Value	Standard Unit	Z-Sco
* Total Leucocytes (WBC) count	ERBA H 560	3.73	/c.mm	

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry
 * : Not considered for evaluation.

Problem Classification: All parameters are acceptable. Only TLC shows unacceptable results. On analysis found that Results entered wrongly. It is a typing error only. otherwise results are ok.

Corrective Action: Before submitting results please check values. Enter @ and Units Carefully. Training for EQAS only done by Metropolis EQAS Team.

Reviewed by: Ashwani

Dated: 14.02.2024

Analyte	Instrument	Result Value	Standard Un
X Neutrophil.	-	<40%	%
# Lymphocyte.	-	>40%	%
X Eosinophil.	-	6-40%	%

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Problem Classification: DLC shows Unacceptable results.
On analysis found that there is mistake on results entry.

Corrective Action: Training for results entry Completed by Metropolis EQAS Plus.

Reviewed by: Ashwini

Dated: 14.2.24