



TECHNICAL CALIBRATION REPORT

1. Name of Customer : Payas pathology Laboratory, Dehu ,Pune
2. Name of instrument : EM200
3. Sr. No. of instrument : v2001070
4. Range of the Photometer : Optical density Linear from 0-2.5
5. Date of Calibration : 16 May 2023

Fully Automatic Biochemistry Auto- Analyzer Parameter check

Parameters	Range	Measured value	
Pressure indication	0.8—1.1 bar	0.9	
Cuvettes blank values	0.02—0.2 O.D.	0.04 to 0 .09	
RCT Temp	37 deg C (+/- 0.5 deg C)	37 deg C	
RGT Temp	8+/- 4degree	11 deg C	
Lamp Voltage	11.9 -12.1 Volts	12.05 volts	
Sr.No.	Wavelengths	Pre Channel gains	Range
1	340	533	300 to 900
2	405	544	----- -----
3	505	635	----- -----
4	546	671	----- -----
5	578	710	----- -----
6	600	640	----- -----
7	660	510	----- -----
8	700	533	----- -----

Next calibration due is 16th May 2024

For TRANSASIA BIO-MEDICALS LTD.

Rajesh Wankhedkar
Regional Service Manager

