

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **Micros ES60** bearing serial number: **110ESOH18819** installed at **KRSNAA Diagnostics, Rajpura, Punjab** has calibrated on 02 April 2024.

Calibrator : ABX MINOCAL
Lot No. : CX493
Expiry Date: 05th May 2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **1 April 2025**.

A circular stamp with the text "HORIBA INDIA PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**