



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 01

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
# Bilirubin Total	Diasys 200 Pro	0.10	mg/dl	0
! Uric Acid	Diasys 200 Pro	2.48	mg/dl	-2.99
! Sodium	ACCULYTE-3P	97.42	mmol/L	-2.04

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
⊙ : Delayed Result Entry  
\* : Not considered for evaluation.

Problem Classification: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	23.20	23.05	0.48	-0.07
@ Creatinine	mg/dl	0.84	0.79	0.81	0.03
# Bilirubin Total	mg/dl	0.10			0.40
@ Total Protein	g/dL	5.76	5.55	0.40	0.07
@ Albumin	g/dL	3.50	3.62	0.47	-4.93
! Uric Acid	mg/dl	2.48	3.62	-2.99	-1.66
@ Calcium	mg/dl	7.12	6.84	0.52	8.15
@ Cholesterol	mg/dl	124.00	130.74	-0.70	-0.63
@ Triglycerides	mg/dl	128.00	122.35	0.31	-0.68
@ HDL	mg/dl	33.00	31.40	1.03	0.23
@ LDL	mg/dl	81.73	72.09	0.59	0.08
@ ALT	U/L	10.40	11.71	-0.49	-0.09
@ AST	U/L	16.41	16.30	0.09	-0.23
@ ALP	U/L	40.44	42.48	-0.54	-0.09

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	86.83	83.97	0.82	0.68
! Sodium	mmol/L	97.42	106.81	-2.04	-1.46
@ Potassium	mmol/L	3.54	3.69	-0.95	-0.77

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

# : Not Evaluated

Ⓢ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	1
Evaluated Parameters	16
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070