

Analyte	Instrument	Result Value	Standard Unit	Z-Score
# Bilirubin Total	Diasys 200 Pro	0.10	mg/dl	0
! Uric Acid	Diasys 200 Pro	2.48	mg/dl	-2.99
! Sodium	ACCULYTE-3P	97.42	mmol/L	-2.04
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	23.20	23.05	0.48	-0.07
@ Creatinine	mg/dl	0.84	0.79	0.81	0.03
# Bilirubin Total	mg/dl	0.10			0.40
@ Total Protein	g/dL	5.76	5.55	0.40	0.07
@ Albumin	g/dL	3.50	3.62	0.47	-4.93
! Uric Acid	mg/dl	2.48	3.62	-2.99	-1.66
@ Calcium	mg/dl	7.12	6.84	0.52	8.15
@ Cholesterol	mg/dl	124.00	130.74	-0.70	-0.63
@ Triglycerides	mg/dl	128.00	122.35	0.31	-0.68
@ HDL	mg/dl	33.00	31.40	1.03	0.23
@ LDL	mg/dl	81.73	72.09	0.59	0.08
@ ALT	U/L	10.40	11.71	-0.49	-0.09
@ AST	U/L	16.41	16.30	0.09	-0.23
@ ALP	U/L	40.44	42.48	-0.54	-0.09

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	86.83	83.97	0.82	0.68
! Sodium	mmol/L	97.42	106.81	-2.04	-1.46
@ Potassium	mmol/L	3.54	3.69	-0.95	-0.77

Legend @ : Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score ≥ 3 - Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	1
Evaluated Parameters	16
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiro Road, Kurla (W),  
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Uric Acid	Diasys 200 Pro	2.54	mg/dl	-3.22
! Sodium	ACCULYTE-3P	100.75	mmol/L	-2.19
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: \_\_\_\_\_

Corrective Action: \_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	69.80	72.36	-0.10	-0.10
@ Urea	mg/dl	23.50	22.82	0.61	-0.07
@ Creatinine	mg/dl	0.92	0.84	0.88	0.03
@ Bilirubin Total	mg/dl	0.40	0.61	-0.86	0.40
@ Bilirubin Direct	mg/dl	0.20	0.14	1.74	1.40
@ Total Protein	g/dL	6.45	6.04	0.79	0.07
@ Albumin	g/dL	3.60	3.66	0.48	-4.93
X Uric Acid	mg/dl	2.54	3.85	-3.22	-1.66
@ Calcium	mg/dl	5.86	6.18	-0.91	8.15
@ Cholesterol	mg/dl	123.00	126.61	-0.59	-0.63
@ Triglycerides	mg/dl	141.00	144.28	-0.47	-0.68
@ LDL	mg/dl	86.95	77.25	0.54	0.08
@ ALP	U/L	36.02	41.55	-0.96	-0.09

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	90.15	87.53	0.84	0.68
! Sodium	mmol/L	100.75	108.73	-2.19	-1.46
@ Potassium	mmol/L	3.73	3.88	-1.15	-0.77

Legend @ : Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score ≥ 3 - Unacceptable  
# : Not Evaluated  
⊙ : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	16
Not Evaluated Parameters	0
Evaluated Parameters	16
Outlier Parameters (X)	1
EQAS Score Biochemistry	93.75 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirod Road, Kurla (W),  
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit	Z-Score
# HDL	Diasys 200 Pro	47.00	mg/dl	0
I Sodium	ACCULYTE-3P	126.46	mmol/L	-2.87
Legend @ : z score $\leq$ 2.0 - Acceptable I : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
# HDL	mg/dl	47.00			0.23
@ ALT	U/L	23.80	35.91	-0.88	-0.09
@ AST	U/L	39.42	80.82	-1.08	-0.23

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	126.46	126.65	0.36	0.68
! Sodium	mmol/L	126.46	157.93	-2.87	-1.46
@ Potassium	mmol/L	6.15	6.84	-1.25	-0.77

Legend @ : Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score ≥ 3 - Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	6
Not Evaluated Parameters	1
Evaluated Parameters	5
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070