

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry ★ : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba ABX Micros ES 60

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.02	2.03	-0.08	0.06
@ Total Leucocytes (WBC) count	/c.mm	3500.00	3510.00	0.00	1.28
@ Haemoglobin	g/dL	5.70	5.65	0.88	0.81
@ MCV (Mean Corpuscular Volume)	fL	118.00	117.78	0.05	-0.36
@ MCH (Mean Corpuscular Hb)	pg	28.20	28.05	0.88	0.55
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	24.00	23.99	0.54	0.71
@ Platelet Count	1000/μL	107.00	119.50	-1.03	-0.55
@ Hematocrit	%	23.70	23.06	-0.08	-0.20
@ RDW CV%	%	17.60	18.04	-0.89	0.57

Legend @ : Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score ≥ 3 - Unacceptable  
# : Not Evaluated  
⊙ : Delayed Result Entry  
★ : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : $2.0 < z$ score $<$ 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry ★ : Not considered for evaluation.				

Problem Classification: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Corrective Action: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Reviewed by: \_\_\_\_\_ Dated: \_\_\_\_\_

Instrument : Horiba ABX Micros ES 60

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.12	2.14	-0.31	0.06
@ Total Leucocytes (WBC) count	/c.mm	7300.00	7640.00	1.35	1.28
@ Haemoglobin	g/dL	6.50	6.48	0.71	0.81
@ MCV (Mean Corpuscular Volume)	fL	123.00	124.00	-0.09	-0.36
@ MCH (Mean Corpuscular Hb)	pg	30.80	30.19	1.06	0.55
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	25.10	24.52	0.77	0.71
@ Platelet Count	1000/ $\mu$ L	203.00	221.70	-1.02	-0.55
@ Hematocrit	%	26.00	26.23	-0.30	-0.20
@ RDW CV%	%	19.00	18.88	0.33	0.57

Legend @ : Acceptable

I : 2.0 < z score < 3.0 - Warning

X : z score  $\geq$  3 - Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kuria (W),  
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable			
X : Unacceptable			
# : Not Evaluated			
⊙ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	>70%	>70%
@ Lymphocyte.	-	<20%	<20%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 O : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	0
Evaluated Parameters	3
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %

*DK Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070



**Metropolis Healthcare Ltd., EQAS**

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

**Outlier And Analyte Summary Report**

Outlier Details For Cycle No 230412 and Sample No 02

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ Delayed Result Entry			

Problem Classification: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Corrective Action: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Diagnosis (Slide)	-	Neutrophilic Leucocytosis	Neutrophilic Leucocytosis
Legend @ : Acceptable		Total Parameters	1
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	1
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Haematology- Peripheral smear diagnosis	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070