



TESTNCL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)

Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)
Mob. +91-8882056269, 9990730550, E-mail-testncl@gmail.com; Website - www.tnclab.com



CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1 Certificate No.: UCCSCC20221207-015 ULR No.: CC328722000005405F
Group: Acceleration & Speed Discipline: Mechanical

Customer's Name & Address: M/s. X-TRA CARE DIAGNOSTICS CHEMBUR
Page No.: 1 of 1 Calibration Performed: At Site
Date of Receipt: 13-02-2024
Service Request No.: UCCSSRN20221207-004
Certificate Issue Date: 14-02-2024
Date of Calibration: 14-02-2024
Next Due Date (Suggested by Customer): 13-02-2025

Equipment Description (UUC) & Identification

Name of the Instrument: Digital RPM Indicator (Centrifuge Machine) Make: Remi
Range: Up to 3000 RPM Model: r-12C Plus
Least Count: 10 RPM Sr. No.: ZHEN-15546
Condition of Instrument: Ok I.D. No.: -----
Location: LAB

Master Equipment Used

Sr. No.	Instrument Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Tachometer	2.5 To 99999 RPM & UCCS/CAL/014	Asian Technology	AT22000000095	06-03-2024


Environmental Conditions: Temperature: 25 ± 5 °C, Humidity: 50 ± 10 %RH
Calibration Procedure: UCCS 03 SOP 033
Calibration Reference Standard: ----

CALIBRATION RESULTS

Sr. No.	Set UUC Value (RPM)	Observed Ref. Standard Value (RPM)	Expanded Uncertainty (±)
1	500	5004	3 % of rdg
2	1000	1007	
3	1500	1503	
4	2000	2008	
5	2500	2506	
6	3000	3007	

End of Certificate

- Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
21. Total liability of our Organization is limited to the invoiced amount.
22. Calibration Certificate without signature are not valid.
23. All our certificates are metrological traceable.
24. The results relate only to the item calibrated.
25. UUC: Unit Under Calibration.
26. Expanded Uncertainty at 95 % confidence level with coverage factor K=2.



Calibrated By

(Amit Kumar)
CEO

Issued & Approved By
Plot No. 113, Pocket A2, Sector
Dwarka, New Delhi-110075