



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :-	Page No.	:- 1 of 1
PAYAS PATHOLOGY LABORATORY Dehu Gaon,Pune	SRF No	:- GTS/240216/05
	Certificate No.	:- GTS/240216/05-001
	Date of Received	:- 16.02.2024
	Date of Calibration	:- 16.02.2024
	Next Calibration Due On	:- 15.02.2025
Ambient Temp. (°C) :- 23.6	Issue Date	:- 21.02.2024
Relative Humidity (%RH) :- 52	Calibration method No.	:- MECH-WI-06
Barometric Pressure (mbar) :- 944.5	ULR No	:- CC295724000005548F
Location of calibration :- In Lab		
Condition of Item :- Ok		

2. Description of Item				
Name	:- Micropipette	Range	:-	100 to 1000 µl
Id No	:- MP/PPL/01	Least Count	:-	5 µl
Make	:- Merilette	Location	:-	Lab
Type	:- Variable	Dept.	:-	Pathology
		Sr No	:-	MD55047/21

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI2023/05/0163	Nishitronics Instrumentation	GTS/WB-01	19.05.2024

*Mechanical Calibration

4.Calibration Results :-				
Calibration Points µl	Standard Reading µl	Set Value on UUC µl	Error in µl	Expanded Uncertainty in ± µl
100	99.6275	100	0.3725	4.30
500	498.0409	500	1.9591	4.30
1000	996.0069	1000	3.9931	4.30

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

V.T.
Calibration Engineer
Varsha T.



Approved By

SwB
Technical Manager
Swapnil Bhagwat

RF-51/00

End of Certificate

Email : globaltechnical007@gmail.com





CALIBRATION CERTIFICATE

1.CUSTOMER	:-	Page No.	:- 1 of 1
PAYAS PATHOLOGY LABORATORY		SRF No	:- GTS/240216/05
Dehu Gaon,Pune		Certificate No.	:- GTS/240216/05-002
		Date of Received	:- 16.02.2024
		Date of Calibration	:- 16.02.2024
		Next Calibration Due On	:- 15.02.2025
		Issue Date	:- 21.02.2024
Ambient Temp. (°C)	:- 23.6	Calibration method No.	:- MECH-WI-06
Relative Humidity (%RH)	:- 53	ULR No	:- CC295724000005549F
Barometric Pressure (mbar)	:- 943.5		
Location of calibration	:- In Lab		
Condition of Item	:- Ok		

2. Description of Item

Name	:- Micropipette	Range	:- 5 to 50 µl
Id No	:- MP/PPL/02	Least Count	:- 5 µl
Make	:- Merilette	Location	:- Lab
Type	:- Variable	Dept.	:- Pathology
		Sr No	:- 22205555

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI2023/05/0163	Nishitronics Instrumentation	GTS/WB-01	19.05.2024

*Mechanical Calibration

4.Calibration Results :-				
Calibration Points µl	Standard Reading µl	Set Value on UUC µl	Error in µl	Expanded Uncertainty in ± µl
10	9.9608	10	0.0392	4.30
25	24.9062	25	0.0938	4.30
50	49.8011	50	0.1989	4.30

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

V.T.
Calibration Engineer
Varsha T.



Approved By

S.B.
Technical Manager
Swapnil Bhagwat

RF-51/00

End of Certificate





GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER		Page No.	:- 1 of 1	
PAYAS PATHOLOGY LABORATORY		SRF No	:- GTS/240216/05	
Dehu Gaon,Pune		Certificate No.	:- GTS/240216/05-003	
Ambient Temp. (°C)		Date of Received	:- 16.02.2024	
Relative Humidity (%RH)		Date of Calibration	:- 16.02.2024	
Barometric Pressure (mbar)		Next Calibration Due On	:- 15.02.2025	
Location of calibration		Issue Date	:- 21.02.2024	
Condition of Item		Calibration method No.	:- MECH-WI-06	
		ULR No	:- CC295724000005550F	

2. Description of Item					
Name	:- Micropipette		Range	:- 100 to 1000 µl	
Id No	:- MP/PPL/03		Least Count	:- 5 µl	
Make	:- Merilette		Location	:- Lab	
Type	:- Variable		Dept.	:- Pathology	
			Sr No	:- 19218226	

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI2023/05/0163	Nishitronics Instrumentation	GTS/WB-01	19.05.2024

*Mechanical Calibration				
4.Calibration Results				
Calibration Points µl	Standard Reading µl	Set Value on UUC µl	Error in µl	Expanded Uncertainty in ± µl
100	99.6491	100	0.3509	4.30
500	498.1441	500	1.8559	4.30
1000	996.0117	1000	3.9883	4.30

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By Calibration Engineer Varsha T.		Approved By Technical Manager Swapnil Bhagwat
--	--	---

RF-51/00 End of Certificate





GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER	:-	Page No.	:- 1 of 1
PAYAS PATHOLOGY LABORATORY		SRF No	:- GTS/240216/05
Dehu Gaon,Pune		Certificate No.	:- GTS/240216/05-004
		Date of Received	:- 16.02.2024
		Date of Calibration	:- 16.02.2024
		Next Calibration Due On	:- 15.02.2025
		Issue Date	:- 21.02.2024
Ambient Temp. (°C)	:- 23.6	Calibration method No.	:- MECH-WI-06
Relative Humidity (%RH)	:- 51	ULR No	:- CC295724000005551F
Barometric Pressure (mbar)	:- 943.3		
Location of calibration	:- In Lab		
Condition of Item	:- Ok		

2. Description of Item

Name	:- Micropipette	Range	:- 5 to 50 µl
Id No	:- MP/PPL/04	Least Count	:- 5 µl
Make	:- pipett man	Location	:- Lab
Type	:- Variable	Dept.	:- Pathology
		Sr No	:- 43337

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI2023/05/0163	Nishitronics Instrumentation	GTS/WB-01	19.05.2024

*Mechanical Calibration

4.Calibration Results

Calibration Points µl	Standard Reading µl	Set Value on UUC µl	Error in µl	Expanded Uncertainty in ± µl
10	9.9605	10	0.0395	4.30
25	24.9029	25	0.0971	4.30
50	49.8224	50	0.1776	4.30

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

Calibration Engineer
Varsha T.



Approved By

Technical Manager
Swapnil Bhagwat

RF-51/00

End of Certificate

