



## PASTEUR DIAGNOSTICS & INSTRUMENTS

F-82, Ind. Area, Phase VII, Mohali-160055, Ph. : 0172-5097009

Email : [pasteurdiag@gmail.com](mailto:pasteurdiag@gmail.com)

GSTIN/UIN : 03ALLPP0756A1ZH

### Certificate of Calibration

*Customer Name* :- *Polo Labs (Inside Ivy Hospital) Hoshiarpur*  
*Instruments* :- *Hematology Analyzer*  
*Model* :- *PCi-5 diff*  
*Serial No* :- *0820207281936*  
*Calibration Date* :- *13/06/2024*  
*Calibration Due Date* :- *12/12/2024*

*This is to certify that we have calibrated the Hematology Analyzer Bearing SR. No - 0820207281936 and on Inspection of the instrument it is observed that the results are within the range and instrument is working satisfactory.*

*Thanking you,*

**For M/S Pasteur Diagnostics & Instruments**

**For Pasteur Diagnostics & Instruments**

**Authorized Signatory**

**Authorized Signatory**



# PASTEUR DIAGNOSTICS & INSTRUMENTS

F-82, Ind. Area, Phase VII, Mahali-160055, Ph. : 0172-5097009

Email : [pasteurdiag@gmail.com](mailto:pasteurdiag@gmail.com)

GSTIN/IN : 03ALLPP0756A1ZH

## Performance Qualification:-

LOW CONTROL					
S/N	PARAMETER	Actual	Target	Range	Remark
1	WBC	3.48	3.54	+ 0.5	PASS
2	RBC	2.08	2.12	± 0.18	PASS
3	HGB	5.8	5.7	± 0.40	PASS
4	MCV	82.3	78.9	± 5.00	PASS
5	PLT	54	46	± 20.00	PASS
NORMAL CONTROL					
S/N	PARAMETER	Actual	Target	Range	Remark
1	WBC	8.54	8.42	+ 1.0	PASS
2	RBC	4.35	4.38	± 0.24	PASS
3	HGB	13.1	13.2	± 0.6	PASS
4	MCV	91.8	89.4	± 5.00	PASS
5	PLT	240	256	+ 40.00	PASS
HIGH CONTROL					
S/N	PARAMETER	Actual	Target	Range	Remark
1	WBC	19.58	19.22	+ 2.50	PASS
2	RBC	5.27	5.19	+ 0.30	PASS
3	HGB	17.3	16.9	+ 0.80	PASS
4	MCV	96.6	95.7	± 5.00	PASS
5	PLT	564	572	± 60.00	PASS