



TESTNCL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)

Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)
Mob. +91-8882056269, 9990730550, E-mail-testncl@gmail.com; Website - www.tnclab.com



CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1	Certificate No.: UCCSCC20221207-015	ULR No.: CC328722000005405F
Group: Acceleration & Speed		Discipline: Mechanical
Customer's Name & Address		Page No.: 1 of 1 Calibration Performed: At Site
Dr. Anita Ranjan Multispecialty Hospital, Farrukhabad Road, Chhibramau, Kannauj		Date of Receipt: 05-03-2024
		Service Request No.: UCCSSRN20221207-004
		Certificate Issue Date: 06-03-2024
		Date of Calibration: 06-03-2024
		Next Due Date (Suggested by Customer): 05-03-2025

Equipment Description (UUC) & Identification					
Name of the Instrument: (Centrifuge Machine)			Make: Remi		
			Model: r-12C Plus		
Range: Up to 2500 RPM			Sr. No.: ----		
Least Count: 10 RPM			I.D. No.: -----		
Condition of Instrument: Ok			Location: LAB		
Master Equipment Used					
Sr. No.	Instrument Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Tachometer	2.5 To 99999 RPM & UCCS/CAL/014	Asian Technology	AT22000000095	06-03-2024
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 25 ± 5 °C, Humidity: 50 ± 10 %RH			UCCS 03 SOP 033	----	


CALIBRATION RESULTS

Sr. No.	Set UUC Value (RPM)	Observed Ref. Standard Value (RPM)	Expanded Uncertainty (±)
1	500	5004	3 % of rdg
2	1000	1008	
3	1500	1012	
4	2000	2016	
5	2500	2520	

End of Certificate

- Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
 2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 3. Total liability of our Organization is limited to the invoiced amount.
 4. Calibration Certificate without signature are not valid.
 5. All our certificates are metrological traceable.
 6. The results relate only to the item calibrated.
 7. UUC: Unit Under Calibration.
 8. Expanded Uncertainty at 95 % confidence level with coverage factor K=2.

Calibrated By



(Signature)
 (Amit Kumar)
 CEO
 Issued & Approved By
 Plot No. 113, Pocket A2, Sector-17,
 Dwarka, New Delhi-110075