



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 4338

Distribution No.: 163-K

Month/Year: April/2024

Instrument ID: SYSMEX

Model Name.: XQ320

Serial No.: 12316

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue &amp; status of the report: 19-06-2024[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.55	4.48	9.03	8.55	0.033	0.57	0.07	0.1	0.006	-0.34
RBC x10 <sup>6</sup> /µl	1	4.79	4.71	9.5	9.1	0.010	1.64	0.08	0.04	0.003	0.77
Hb g/dl	1	13.1	13.1	26.2	25.5	0.030	0.86	0	0.1	0.007	-1.35
HCT%	1	40	39.4	79.4	79.2	0.170	0.05	0.6	0.4	0.023	0.54
MCV-fl	1	83.7	83.5	167.2	174	0.314	-0.83	0.2	0.3	0.022	-0.27
MCH-Pg	1	27.8	27.3	55.1	56.2	0.067	-0.68	0.5	0.2	0.016	1.35
MCHC-g/dl	1	33.2	32.8	66	64.35	0.161	0.39	0.4	0.3	0.019	0.34
Plt. x10 <sup>3</sup> /µl	1	77	76	153	179.5	1.782	-0.54	1	5	0.297	-0.90
Retic %	2	18	16	34	16	0.260	2.53	2	0.5	0.036	2.53

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=00 , Poly=11 L=85, E=01, Mono/Promono=03 , B1=00 P.M.=00, Mye=00, Meta=00, Other=00	Lymp: 77-88, Poly: 8-14.25, mono: 1-3, nRBC/Blast/Myelo/Meta/Eosino: 0-5		
RBC Morphology	3	Predominantly Normochromic Normocytic	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Microcytic		
Diagnosis	3	Lymphoproliferative disorder favouring chronic Lymphocytic Leukaemia(CLL)	Chronic Lymphoproliferative Disorder/CLL		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /μl	1	303	303	79.21	89.44	5.61	5.28	15.18	5.28
RBC x10 <sup>6</sup> /μl	1	303	303	89.44	90.76	5.28	3.96	5.28	5.28
Hb g/dl	1	303	303	90.1	91.42	5.61	3.96	4.29	4.62
HCT%	1	303	302	93.05	89.4	4.97	5.3	1.98	5.3
MCV-fl	1	303	302	92.05	85.1	5.3	7.28	2.65	7.62
MCH-Pg	1	303	302	87.09	92.38	6.95	4.64	5.96	2.98
MCHC-g/dl	1	303	302	93.05	93.38	3.64	3.31	3.31	3.31
Plt. x10 <sup>3</sup> /μl	1	303	302	93.71	91.06	3.64	5.63	2.65	3.31
ReticCount%	2	303	251	95.22	86.85	3.19	7.17	1.59	5.98
PS Assessment	3	303	249	Satisfactory :92.74%, Borderline Sat. :3.30%, Unsatisfactory :3.96%					

**Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme


**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaimseqap.com](http://www.ishtmaimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----