



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.09	2.03	0.63	0.25
@ Total Leucocytes (WBC) count	/c.mm	2150.00	2047.65	0.16	-0.07
@ Haemoglobin	g/dL	5.40	5.47	-0.37	-0.49
@ MCV (Mean Corpuscular Volume)	fL	114.20	117.12	-0.39	-0.26
@ MCH (Mean Corpuscular Hb)	pg	26.00	26.96	-0.72	-0.57
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	22.80	23.00	-0.18	-0.24
@ Platelet Count	1000/ μ L	136.00	130.07	0.82	0.26
@ Hematocrit	%	23.90	23.76	0.03	-0.01
@ RDW CV%	%	15.90	19.25	-1.89	-1.40

Legend @ : Acceptable

! : $2.0 < z \text{ score} < 3.0$ - Warning


X : $z \text{ score} \geq 3$ - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070