



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: BALAJEE Diagnostics. Near Bohar Police Chowki, Ring Road, Dubagga, Lucknow.	Service request No. & date	A/02/20/02/24&20/02/2024
		ULR no.	CC223924000001028F
		Cert. No.	AT24000001028
		Date of Receipt of DUC	20/02/2024
		Date of calibration	20/02/2024
		Date of issue	21/02/2024
		Suggested due date	19/02/2025

Instrument Details			
Instrument name	Mini Centrifuge	Location	-----
Make / Model No.	Remi / R-303	Accuracy	-----
Range / Size	0 to 3500 rpm	Visual Inspection	OK
Least Count	N/A	Calibration Performed at	Site
I.D. No. / Sr. No.	BD/RF-01 / ZIDN-16351		

Detail of reference standards & Major equipment's used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/A.C92792		
Certificate No.	QCTS/00053/10		
Calibration Validity	19/10/2024		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Serial No.	Position	Calibration Results	
		Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	Position: 1	1135	±0.2 %rdg.
02.	Position: 2	1590	
03.	Position: 3	2205	
04.	Position: 4	2812	
05.	Position: 5	3488	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Signature)
Calibration Engg./TM
(M. K. PAL)
Form No.- QF-47



Approved By
(Signature)
(Q.M./TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of report---