



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2024



PC-1024

Lab Name **CLINICA DIAGNOSTICS** Lab No **20244**  
 Constituent Group **Chemistry II** Date of Result Entered : **17/03/2024**  
 PT item **Lyophilized human serum based** Date of Report Published : **06/04/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1019	176.25	5.91	10.41	175 mg/dL	-0.12	0.65
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1028	25.73	10.25	2.64	26.4 mg/dL	0.25	0.16
3	CREATININE II	Enzymatic Colorimetric II	Any Analyser (Automation / Semi Automation)	550	0.91	10.61	0.10	0.89 mg/dL	-0.21	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	995	0.76	16.38	0.12	0.69 mg/dL	-0.56	0.01
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1071	5.72	7.33	0.42	6.44 g/dL	1.72	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1051	3.50	4.80	0.17	3.65 g/dL	0.89	0.01
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1001	3.23	10.29	0.33	3.5 mg/dL	0.81	0.02
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1071	111.23	8.30	9.24	110 mg/dL	-0.13	0.56
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	853	78.55	11.83	9.29	80.2 mg/dL	0.18	0.64

u\* - Method of Uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

## LAB ADDRESS :

CLINICA DIAGNOSTICS

A UNIT OF CLINICA LIFE CARE PVT LTD, 886/B/1, KRISHNANAGAR ROAD (NG-34)

BARASAT

WEST BENGAL 700124

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***