



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2024



PC-1024

Lab Name	CLINICA DIAGNOSTICS	Lab No	20244
Constituent Group	Chemistry II	Date of Result Entered :	17/03/2024
PT item	Lyophilized human serum based	Date of Report Published :	06/04/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1019	176.25	5.91	10.41	175 mg/dL	-0.12	0.65
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1028	25.73	10.25	2.64	26.4 mg/dL	0.25	0.16
3	CREATININE II	Enzymatic Colorimetric II	Any Analyser (Automation / Semi Automation)	550	0.91	10.61	0.10	0.89 mg/dL	-0.21	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	995	0.76	16.38	0.12	0.69 mg/dL	-0.56	0.01
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1071	5.72	7.33	0.42	6.44 g/dL	1.72	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1051	3.50	4.80	0.17	3.65 g/dL	0.89	0.01
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1001	3.23	10.29	0.33	3.5 mg/dL	0.81	0.02
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1071	111.23	8.30	9.24	110 mg/dL	-0.13	0.56
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	853	78.55	11.83	9.29	80.2 mg/dL	0.18	0.64

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

CLINICA DIAGNOSTICS
A UNIT OF CLINICA LIFE CARE PVT LTD, 886/B/1, KRISHNANAGAR ROAD (NG-34)
BARASAT
WEST BENGAL 700124

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

6/25/24, 3:37 PM

External Quality Assurance Scheme - Print Monthly Summary

**Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *******