



## HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,  
New Delhi 110020, India  
Tel : +91 (11) 4646 5000 / 4669 5001  
Fax : +91 (11) 4646 5020  
http : //www.horiba.com  
CIN : U73100DL2006PTC153232

HIN/MED/2023-2024/50045245  
05aug2024

### CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **YUMIZEN H500** bearing serial number: **007YOXH03585** installed at **tapdiya diagnostic Aurangabad** as calibrated on 04<sup>th</sup>aug 2024

Calibrator

:ABXMINOCAL

LotNo. :CX494

ExpiryDate:07<sup>th</sup>sep24

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 03<sup>th</sup>aug 25.



ShrishDixit

(Head-Products &Marketing)For**HoribaIndiaPvt**  
**.Ltd.**

# HORIBA

Medical

## **HORIBAIndiaPrivateLimited**

246,OkhlaIndustrialEstatePhase-  
III,NewDelhi110020,IndiaTel:+91(11)46  
465000/46695001Fax:+91(11)4646502  
0

<http://www.horiba.com>CIN:U731  
00DL2006FTC153232