

Analyzer Report

163
Whole Blood-CBC+DIFF
EQAS 01

4/2024 12:16
[only accounts for this test sample]

Para	Result	Unit	Ref Range	Para	Result	Unit	Ref Range	Para	Result	Unit	Ref Range
WBC	8.44	/nL	4.00-10.00	RBC	3.76	10 ¹² /L	3.50-5.50	P-LCC	36	%	30-90
Neu#	4.15	/nL	2.00-7.00	HGB	13.3	g/dL	11.0-16.0	P-LCR	31.4	%	11.0-45.0
Lym#	3.17	/nL	0.80-4.00	HCT	38.7	%	37.0-54.0				
Mon#	0.25	/nL	0.12-1.20	MCH	102.8	fL	80.0-100.0				
Eos#	0.49	/nL	0.02-0.50	MCHC	35.4	g/dL	27.0-34.0				
Bas#	0.38	/nL	0.00-0.10	RDW-CV	34.4	%	32.0-36.0				
Neu%	49.1	%	50.0-70.0	RDW-SD	16.2	fL	11.0-16.0				
Lym%	37.6	%	20.0-40.0	MPV	27.5	fL	35.0-56.0				
Mon%	3.0	%	3.0-12.0	P-L	10.4	fL	7.0-11.0				
Eos%	5.8	%	0.5-5.0	PDW-CV	0.136	%	0.150-0.170				
Bas%	4.5	%	0.0-1.0	PDW-SD	14.8	fL	9.0-17.0				
NLR	1.31			PCT	0.286	%	0.108-0.282				
PLR	86.75										

Result 36 Unit 10⁹/L Ref Range 30-90

Result 31.4 Unit % Ref Range 11.0-45.0

22/04/2024 12:13
[only accounts for this test sample]

PLC - P-11 L 88 B 01

PLs Red Blood Cells seen predominantly are normochromic normochromic in nature with mild anisocytosis and hypochromia. Leukocytes are seen.



Diagnosis: Chronic Lymphocytic Leukemia

Hematology Analyzer Report

Sample ID: 1641
 Code: Whole Blood-CBC+DIFF
 Name: EQAS 02
 Operator: Admin
 Test Time: 22/04/2024 12:16
 [The test result only accounts for this test sample]

Para	Result	Unit	Ref. Range	Para	Result	Unit	Ref. Range
WBC	8.24	/nL	4.00-10.00	RBC	3.79	10 ¹² /L	3.50-5.50
Neu#	3.86	/nL	2.00-7.00	HGB	13.4	g/dL	11.0-16.0
Lym#	3.21	/nL	0.80-4.00	HCT	39.0	%	37.0-54.0
Mon#	0.35	/nL	0.12-1.20	MCV	102.8	fL	80.0-100.0
Eos#	0.52	/nL	0.02-0.50	MCH	35.3	pg	27.0-34.0
Bas#	0.30	/nL	0.00-0.10	MCHC	34.3	g/dL	32.0-36.0
Neu%	47.0	%	50.0-70.0	RDW-CV	16.0	%	11.0-16.0
Lym%	38.9	%	20.0-40.0	RDW-SD	59.1	fL	35.0-56.0
Mon%	4.2	%	3.0-12.0	PLT	285	10 ⁹ /L	100-300
Eos%	6.3	%	0.5-5.0	MPV	10.6	fL	7.0-11.0
Bas%	3.6	%	0.0-1.0	PDW-CV	0.147		0.150-0.170
NLR	1.20			PDW-SD	16.9	fL	9.0-17.0
PLR	88.79			PCT	0.303	%	0.108-0.282

Para	Result	Unit	Ref. Range	LAS	DIFF
P-LCC	93	10 ⁹ /L	30-90		
P-LCR	32.6	%	11.0-45.0		






Hematology Analyzer Report

Sample ID: 163
 Code: Whole Blood-CBC+DIFF
 Name: EQAS 01
 Operator: Admin
 Test Time: 22/04/2024 12:13
 [The test result only accounts for this test sample]

Para	Result	Unit	Ref. Range	Para	Result	Unit	Ref. Range
WBC	8.44	/nL	4.00-10.00	RBC	3.76	10 ¹² /L	3.50-5.50
Neu#	4.15	/nL	2.00-7.00	HGB	13.3	g/dL	11.0-16.0
Lym#	3.17	/nL	0.80-4.00	HCT	38.7	%	37.0-54.0
Mon#	0.25	/nL	0.12-1.20	MCV	102.8	fL	80.0-100.0
Eos#	0.49	/nL	0.02-0.50	MCH	35.4	pg	27.0-34.0
Bas#	0.38	/nL	0.00-0.10	MCHC	34.4	g/dL	32.0-36.0
Neu%	49.1	%	50.0-70.0	RDW-CV	16.2	%	11.0-16.0
Lym%	37.6	%	20.0-40.0	RDW-SD	60.0	fL	35.0-56.0
Mon%	3.0	%	3.0-12.0	PLT	275	10 ⁹ /L	100-300
Eos%	5.8	%	0.5-5.0	MPV	10.4	fL	7.0-11.0
Bas%	4.5	%	0.0-1.0	PDW-CV	0.136		0.150-0.170
NLR	1.31			PDW-SD	14.8	fL	9.0-17.0
PLR	86.75			PCT	0.286	%	0.108-0.282

Para	Result	Unit	Ref. Range	LAS	DIFF
P-LCC	86	10 ⁹ /L	30-90		
P-LCR	31.4	%	11.0-45.0		



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029

Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4484
 Instrument ID: Countcell

Distribution No.: 163-L
 Model Name.: Penta 2.0 Auto
 hematology analyzer

Month/Year: April/2024
 Serial No.: 0840519220422

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmalimseqap.com
 Date of issue & status of the report: 19-06-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /μl	1	8.44	8.24	16.68	16.14	0.046	0.48	0.2	0.1	0.008	0.84
RBC x10 ⁶ /μl	1	3.79	3.76	7.55	7.96	0.009	-1.78	0.03	0.03	0.002	0.00
Hb g/dl	1	13.4	13.3	26.7	26.4	0.028	0.40	0.1	0.1	0.008	0.00
HCT%	1	39	38.7	77.7	83.05	0.206	-0.95	0.3	0.4	0.024	-0.27
MCV-fl	1	102.93	102.9	205.83	207.15	0.423	-0.12	0.03	0.3	0.020	-0.91
MCH-Pg	1	35.37	35.36	70.73	66.25	0.060	0.99	0.01	0.3	0.021	-0.98
MCHC-g/dl	1	34.37	34.36	68.73	63.2	0.152	1.39	0.01	0.3	0.022	-0.98
Plt. x10 ³ /μl	1	285	275	560	509	1.829	1.04	10	6	0.352	0.77
Retic %	2	4.9	4.5	9.4	5.45	0.111	1.30	0.4	0.3	0.022	0.22

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT	
DLC%	3	Nrbcs=0 , Poly=11 L=88, E=00, Mono/Promono=00 , B1=01 P.M.=00, Mye=00, Meta=00, Other=Pp	Lymph: 77-88, Poly: 9-15, mono: 1-3, nRBC/Blast/Myelo/Meta/Eosino: 0-1	
RBC Morphology	3	Red blood cells seen predominantly are normocytic normochromic in nature with mild anisocytosis and hypochromia. Occasional microcytic hypochromic cells are seen	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Microcytic	
Diagnosis	3	Chronic Lymphocytic Leukemia	Chronic Lymphoproliferative Disorder/CLL	

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /μl	1	319	319	83.7	87.46	3.13	5.02	13.17	7.52
RBC x10 ⁶ /μl	1	319	319	86.83	86.52	7.21	5.02	5.96	8.46
Hb g/dl	1	319	319	87.15	85.58	7.84	7.21	5.01	7.21
HCT%	1	319	319	91.22	87.77	4.7	5.64	4.08	6.59
MCV-fl	1	319	319	88.71	85.89	6.9	7.84	4.39	6.27
MCH-Pg	1	319	319	89.66	90.91	4.7	3.76	5.64	5.33
MCHC-g/dl	1	319	319	89.97	90.28	5.64	5.02	4.39	4.7
Plt. x10 ³ /μl	1	319	319	88.4	91.22	6.27	4.7	5.33	4.08
ReticCount%	2	319	210	91.9	95.71	4.76	3.81	3.34	0.48
PS Assessment	3	319	201	Satisfactory :91.55%, Borderline Sat. :4.07%, Unsatisfactory :4.38%					

Comments:

- 1). Among Lab (EQA) : Results acceptable.
- 2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values – Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values – Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between “0 to ±2” are texted in green colour. Z score value between “±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample (x-y) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme


Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme’s website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)
PT Co-ordinator: ISHTM-AIIMS-EQAP
Department of Hematology, AIIMS, New Delhi

-----End Of Report-----