

AGD Biomedicals (P) Ltd.
Address- Mehta Trade Center, Sir M.V.Road ,
Andheri East, Mumbai-400099
Tel – +91-22-28231061/66
Email : sales@agdbio.com

DOC.NO.- CC/AGD 2260/CCNR/02/00
DATE: 29 June 2024

CALIBRATION CERTIFICATE

This is to certify that the AGD 2260 Fully Automated Clinical Chemistry Analyzer bearing serial number: K3-E2203N118L installed at CHETHANA CENTRE FOR NEUROPHYSIATRIC REHABILITATION MALAPARAMBA, CALICUT KERALA. is company Calibrated .

Verified with Quality control material Randox serum Control (Level 2) .
The reports of Control values were all found in acceptable range.
Data given to laboratory.


Instrument is ready for routine work.

Please refer our Service Report on dated 26 June, 2024 for your reference.

Calibration Date: 26 June , 2024
Next Calibration Due on: 25 June, 2025

Enclosed :

- 1) QC Run Data
- 2) Assay Sheet of Control
- 3) Printouts

Authorized Signatory *Amal Mohan .k*




For AGD BIOMEDICALS Pvt. Ltd.