

AGD Biomedicals (P) Ltd.
Address- Mehta Trade Center, Sir M.V.Road ,
Andheri East, Mumbai-400099
Tel – +91-22-28231061/66
Email : sales@agdbio.com

DOC.NO.- CC/EL-120/CCFNR/123/00
DATE: 28 June 2024

CALIBRATION CERTIFICATE

This is to certify that the Electrolyte Analyzer Erma EL120 bearing serial number: **BC2108007** installed at **Chethana Center For Neuropsychiatric Rehabilitation, Providence Collage Road, Malaparamba P.O. Calicut, Kerala** is company Calibrated .

Verified with Quality control material ISE Tritrol (Normal, Low & High) .
The reports of Control values were all found in acceptable range.
Data given to laboratory.

Instrument is ready for routine work.


Please refer our Service Report No.93534 on dated 28 June 2024 for your reference.

Calibration Date: 28 June 2024
Next Calibration Due on: 27 December 2024

Enclosed :

- 1) IQC Run Data
- 2) Assay Sheet of Control
- 3) Printouts

Authorized Signatory

Amal Mohan.k




For AGD BIOMEDICALS Pvt. Ltd.