



CALIBRATION CERTIFICATE

CERTIFICATE No: RTK /21156/2022-23

Customer:

CHETHANA CENTRE FOR
NEUROPSYCHIATRIC REHABILITATION
MALAPARAMBA, KOZHIKODE

Date of Calibration: 30/03/2024
Calibration up to : 29/03/2025

This is to certify that the following item has been calibrated and functionally checked as per our factory norms.
BACTERIOLOGICAL INCUBATOR: RI – 12 SS

| DESCRIPTION | |
|---|---|
| INSTRUMENT: BACTERIOLOGICAL INCUBATOR | Cat.No . RI – 12 SS Product :SL NO:21156 |
| AREA OF USAGE: BACTERIOLOGICAL INCUBATOR | PHYSICAL APPEARANCE: Good Condition |

Calibrated at (RT) **32°C**

| Sl. No. | Calibration Temp: set | Instrument temp: reading | Deviation |
|---------|--------------------------|--------------------------|-----------|
| 1 | 32°C | 34°C | +2 °C |
| 2 | 60°C | 65°C | +5 °C |
| 3 | 95°C | 100°C | +5 °C |

CALIBRATED BY.

Gijomathew

ENGINEER PRODUCTION.

APPROVED BY.

Balakrishnan

INCHARGE.

This report may not be reproduced except in full without the written permission from rotek



ISO 9001:2015



Certificate No.
SCPL2504

