



**CERTIFICATE NO. RTK/21262/2023-24**

**Customer:**

**CHETHANA CENTRE FOR  
NEUROPSYCHIATRIC REHABILITATION  
MALAPARAMBA, KOZHIKODE**

Date of Calibration: 30/03/2024

**calibration up to: 29/03/2025**

This is to certify that the following item has been calibrated and functionally checked as per our factory norms.

DESCRIPTION	
<b>INSTRUMENT: MEDICO Centrifuge</b>	<b>Cat. No. RMC – 6 AL Product :SL NO:21262</b>
<b>AREA OF USAGE:CENTRIFUGE</b>	<b>PHYSICAL APPEARANCE: Good Condition</b>

SPECIFICATIONS	
RPM: Maximum: 5000 RPM	Motor :1/20 HP AC/DC MOTOR  Checked at : 220V AC
Watts      40	AMPS: 0.17

Tachometer Make used for calibration: TECLOCK Corporation, Japan

Knob position	Calibrator RPM read	Instrument RPM read
1	800	900
2	1600	1620
3	2400	2500
4	3200	3200
5	4000	3800

CALIBRATED BY

UNIKRISHNAN

APPROVED BY

BALAKRISHNAN MV



ISO 9001:2015



Certificate No.  
SCPL2504

