

## Urine Routine Analysis

### ALL METHOD REPORT

Cycle - 13/2024

Date: 29/06/2024

Lab Code: 3134

Round No -06

| Parameters         | Units | No.of. Participants | Consensus |      | Consensus Result | Your Value | Remark           |
|--------------------|-------|---------------------|-----------|------|------------------|------------|------------------|
|                    |       |                     | Number    | %    |                  |            |                  |
| Specific Gravity   | -     | 399                 | 372       | 93%  | 1.010 - 1.020    | 1.010      | Within Consensus |
| pH                 | -     | 399                 | 358       | 90%  | 6.0              | 6.0        | Within Consensus |
| Leukocytes         | /μl   | 387                 | 378       | 98%  | Negative         | .          | -                |
| Nitrite            | -     | 381                 | 376       | 99%  | Negative         | .          | -                |
| Protein            | mg/dl | 399                 | 389       | 98%  | Negative         | Negative   | Within Consensus |
| Glucose            | mg/dl | 399                 | 393       | 99%  | Negative         | Negative   | Within Consensus |
| Ketones            | -     | 395                 | 394       | 100% | Negative         | .          | -                |
| Urobilinogen       | mg/dl | 393                 | 385       | 98%  | Negative         | .          | -                |
| Bilirubin          | mg/dl | 393                 | 376       | 96%  | Negative         | .          | -                |
| Blood (Hemoglobin) | /μl   | 395                 | 384       | 97%  | Negative         | Negative   | Within Consensus |

|         |                 |                       |                  |                  |                           |
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|
| Legends | ≥ 60% Consensus | (*) Outside Consensus | (@) No Consensus | (.) Not Reported | (#)Late Result Submission |
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|

Chief Coordinator



Dr. Sanjay Mehrotra

Prepared By: DP

Programme Director



Dr. Bandana Mehrotra

Checked By: 

Doc. No.: ASS / FR / 10A / R 00 / Dt.: 05.01.2024