



R-TECH CALIBRATION

ISO/IEC 17025:2017 - NABL ACCREDITED LABORATORY

52/1, Pathak Para Road, Behala, Room No. G1, Kolkata - 700 060



CALIBRATION CERTIFICATE

| | | | | | |
|-----------------|---------------------|------------------------|-------------------------|--------------------|---------------|
| ULR : | CC33012400000010F | | CERTIFICATE NO : | RTC/24/010 | |
| Date of Receipt | Date of Calibration | Certificate Issue Date | Location of Calibration | Suggested Due Date | SPF No / Date |
| 04-Jan-24 | 04-Jan-24 | 05-Jan-24 | SITE | 03-Jan-25 | SPF/24/003 |

| | | | | | |
|------------------------|--|--|--------------------------|--------------------------------|--|
| Name of Customer : M/s | UNITY GASTRO CARE AND DIAGNOSTIC CENTRE | | Environmental Conditions | | |
| | 17, Thakurpalli Road, Kadamtall, Krishnanagar Agartala, Tripura, Pin 799 001 | | Temperature: 25 ± 4°C | Relative Humidity: 50 ± 20% RH | |

| | | | | | | |
|--------------------|------------|--------------------------|------------|---------|------------|--------------|
| Description of DUC | Centrifuge | | Make | Remi | Range | 3500 RPM |
| Name of Equipment: | Centrifuge | | Model | R-303 | L.C. | 1 PPM |
| Condition of Item: | Good | Calibration Discipline : | Mechanical | ID. No. | N/M | Type Analog |
| Location of DUC : | N/M | Calibration Parameter : | Speed | Sr. No. | ZFGN-23169 | Accuracy N/M |

REFERENCE STANDARD & MEASUREMENT TRACEABILITY

| | | | | | |
|-----------------|------------|------------------|------------------|-----------|------------------|
| Name | Serial No. | Certificate No. | Calibration Date | Due Date | Traceability No. |
| Dig. Tachometer | 1610208981 | ATL/E/220523/009 | 15-May-23 | 15-May-24 | CC-2590 |

Calibration Procedure No.: The Calibration is carried out as per the "RTC-CP-M-04".

CALIBRATION RESULT

| Sr. No. | Calibration Point | Avg. Standard Reading | Avg. DUC Reading | Deviation | Expanded Uncertainty |
|---------|-------------------|-----------------------|------------------|-----------|----------------------|
| | (RPM) | (RPM) | (RPM) | (RPM) | (RPM) |
| 1 | 500 | 500.4 | 500 | -0.4 | ±16 |
| 2 | 1000 | 1001 | 1000 | -1 | ±16 |
| 3 | 2000 | 2003 | 2000 | -3 | ±16 |
| 4 | 3000 | 3002 | 3000 | -2 | ±16 |
| 5 | 3500 | 3501 | 3500 | -1 | ±16 |

Note: *DUC: Device Under Calibration, N/M : Not Mentioned.

The Error is positive except when indicated otherwise. It has to algebraically subtracted from DUC Reading in order to get the standard reading.

The Expanded Uncertainty associated with measurement at 95.4% confidence level & coverage factor K = 2 is above under uncertainty column.

Note :-

1. This certificate refers only to the particular item(s) submitted for calibration at Site/Laboratory.
2. Results reported are valid at the time of and under the stated conditions of measurement.
3. Certificate should not be reproduced, except in full, without the prior permission of Laboratory.
4. Next Calibration due date mentioned as per customer requirement.
5. All reference standard used for calibration are traceable to National Standard.

***** End of Certificate *****

Chandan



Ramy

| | |
|-----------------------------------|--|
| Calibration Engineer (Chandan) | Authorized Signatory (Ramasare Prajapati) |
| Format : RTC-QF-7.8-02 | Page No.: 01/01 |

033-3557 4693

7044226000 / 7891753269

rtchgroups06@gmail.com
lab@rtechcalibration.in
info@rtechcalibration.in

http://rtechcalibration.in

CALIBRATION

SALES & SERVICE

REPAIR