

CALIBRATION CERTIFICATE

| | | | |
|---------------------------|---|---------------------|------------------------|
| Customer Name And Address | M/S : Unity Gastrocare And Diagnostic Centre 17, Thakurpalli Road, Kadamtali, Krishnanagar Agartala, Tripura, Pin 799 001 | Certificate No. | MT/05/030124/AC7/00421 |
| | | ULR No. | CC333324000000417F |
| | | SRF No. & Date | MT/05/03.01.24 |
| | | Receipt Date | 03/01/2024 |
| | | Date of Calibration | 03/01/2024 |
| | | Suggested Due Date | 02/01/2025 |
| | | Date of Issue | 04/01/2024 |

| Instrument Details | | | |
|--------------------|-----------------------------------|--------------------------|---------|
| Instrument name | Pipette | Sr. No. | QW06955 |
| Make /Model No | Thermo Scientific /Finnpipette F3 | Location | ----- |
| Range / Size | 5 – 50 µL | Accuracy | ----- |
| Least Count | 0.5 µL | Visual Inspection | OK |
| I.D. No. | ----- | Calibration Performed At | Lab |

| Detail of Reference Standard & Major Equipment Used | | | |
|---|----------------------------|--|--|
| Equipment Name | Digital Electronic Balance | | |
| Make | Radwag | | |
| Model / SR No. | AS 82/220. R2 / 640970 | | |
| Certificate No. | TYCON/W/10/2023/965 | | |
| Calibration Validity | 27/10/2024 | | |
| Calibration By | Tycon Engineering | | |

| Environmental Condition | Value | Calibration Reference | Calibration Procedure |
|-------------------------|---------|-----------------------|-----------------------|
| Room Temperature | 23.4 °C | NABL129, ISO-8655-6 | |
| Relative Humidity | 54% | | CP/M/M&V/03 |
| Water Temperature | 23.1 °C | | |

Calibration Results

| Serial No. | UUC in (µl) | Std. (Master) Value in (gm) | Std. (Master) Value Converted into (µl) | Uncertainty At 95% C.L. (coverage factor k=2) |
|------------|-------------|-----------------------------|---|---|
| 01. | 20 | 0.02001 | 20.01 | ±0.2µl |
| 02. | 30 | 0.03004 | 30.04 | ±0.2µl |
| 03. | 50 | 0.05006 | 50.06 | ±0.2µl |

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k = 2$ such that the coverage probability corresponds to approximately 95%.
- ❖ (3) The above results are valid at the time of and under the stated conditions measurement.
- ❖ (4) This certificate refers only to the particular item submitted for calibration.
- ❖ (5) Next calibration due date given as requested by the customer.
- ❖ (6) Certificate shall not reproduced expect in full, without the Written Approval of Micro Technology.
- ❖ (7) Coefficient of Cubical Thermal Expansion for material Borosilicate glass 3.3 is $(9.9 \times 10^{-6} / ^\circ\text{C})$.

Calibrated By
(Calibration Engineer)
(Kanhaya)
Format No. F01(7.8)

.....End of the Certificate.....

Approved By
(Quality Manager)
(Amit Saoni)
Page No. 1 of 1