



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233923000003821F	Calibration Done on : 25.12.2023	S.R.F.No. : 0001/25
Issue Date : 28.12.2023	Calibration Valid Upto : 24.12.2024	S.R.F Date : 25.12.2023
UUC Receipt Date : 25.12.2023	Certificate No. : PCS/3821	Page : 01 of 01
Calibrated for M/S. AAKASH HOSPITAL 90/43, Malviya Nagar, New Delhi-110017		

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(10 to 100) µl
Least Count	0.2 µl	Condition of Item	Good
Manufactured By	Dragon Lab	Calibration Performed at	Lab
Manufacturer M/S.No.	YE19BAM0061506	Location	Lab
Party ID Mark	AH/Pipt/02	Ref. Standard / Method	ISO-4787
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 ± 3) °C. Relative Humidity : (50 ± 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Avg. Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (µl)
10.0	10.00	0.00	0.13
50.0	50.11	-0.11	0.13
100.0	100.23	-0.23	0.13

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : GC23392300003822F	Calibration Done on : 25.12.2023	S.R.F.No. : 0001/25
Issue Date : 28.12.2023	Calibration Valid Upto : 24.12.2024	S.R.F Date : 25.12.2023
UUC Receipt Date : 25.12.2023	Certificate No. : PCS/3822	Page : 01 of 01
Calibrated for M/S. AAKASH HOSPITAL 90/43, Malviya Nagar, New Delhi-110017		

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(100 to 1000) µl
Least Count	1 µl	Condition of Item	Good
Manufactured By	Dragon Lab	Calibration Performed at	Lab
Manufacturer M/S.No.	YE22BBC00702	Location	Lab
Party ID Mark	AH/Pipt/03	Ref. Standard / Method	ISO-4787
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 ± 3) °C. Relative Humidity : (50 ± 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Avg. Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (µl)
100	100.05	-0.05	0.13
500	500.35	-0.35	0.56
1000	1000.63	-0.63	0.56

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD

DY. (QM/T)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233923000003823F	Calibration Done on : 25.12.2023	S.R.F.No. : 0001/25
Issue Date : 28.12.2023	Calibration Valid Upto : 24.12.2024	S.R.F Date : 25.12.2023
UUC Receipt Date : 25.12.2023	Certificate No. : PCS/3823	Page : 01 of 01

Calibrated for M/S.

AAKASH HOSPITAL
90/43, Malviya Nagar, New Delhi-110017

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(5 to 50) µl
Least Count	0.5 µl	Condition of Item	Good
Manufactured By	Dragon Lab	Calibration Performed at	Lab
Manufacturer M/S No.	YE228AX044982	Location	Lab
Party ID Mark	AH/Pip/04	Ref. Standard / Method	ISO-4787
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 ± 3) °C. Relative Humidity : (50 ± 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Avg. Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (µl)
10.0	10.06	-0.06	0.13
30.0	29.91	0.09	0.13
50.0	50.16	-0.16	0.13

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf