



# Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092  
Ph.: 011-42657237 Mob.: 9810804449, 8368671530  
E- mail: primac21@gmail.com



Format No.  
F01(7.8/01)

## CALIBRATION CERTIFICATE

ULR No. : CC23392300003894F	Calibration Done on : 26.12.2023	S.R.F.No. : 0002/26
Issue Date : 28.12.2023	Calibration Valid Upto : 25.12.2024	S.R.F Date : 26.12.2023
UUC Receipt Date : N/A	Certificate No. : PCS/3894	Page : 01 of 01
Calibrated for M/S. <b>AAKASH HOSPITAL</b> 90/43, Malviya Nagar, New Delhi-110017		

### Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	(0 -6000 ) RPM
Least Count	10 RPM	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr. No.	NEYA4 / ZFEN-21303	Location	Lab
Party ID Mark	AH/CF/02	Ref. Standard / Method	IS: 12508
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-S-04

Environmental conditions (s) : Temperature : (25 ± 10) °C. Relative Humidity : (50 ± 15) % RH

### Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	PCS/DTMC/01 / 109-0595Y	Transcal	TSC/22-23/18608-1	14.03.2024

### Calibration Results:

UUC Value (RPM)	Avg. Std. Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
500	500.6	-0.6	1.65%
1000	1000.9	-0.9	1.60%
2000	2001.3	-1.3	1.60%
3000	3001.6	-1.6	1.60%
5000	5002.3	-2.3	1.60%

Remark :-

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD  
DY.(QM/TM)

\*\*\*\*\* END OF CERTIFICATE \*\*\*\*\*



Note : The Certificate has issued subject to conditions stated back side overleaf





# Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



CC-2339

Format No.  
F01(7.8/01)

## CALIBRATION CERTIFICATE

ULR No. : GC233923000003895F	Calibration Done on : 26.12.2023	S.R.F.No. : 0002/26
Issue Date : 28.12.2023	Calibration Valid Upto : 25.12.2024	S.R.F Date : 26.12.2023
UUC Receipt Date : N/A	Certificate No. : PCS/3895	Page : 01 of 01
Calibrated for M/S. <b>AAKASH HOSPITAL</b> 90/43, Malviya Nagar, New Delhi-110017		

### Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	(0 - 3000 ) RPM
Least Count	1/10 RPM	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr. No.	N/A	Location	Lab(Aakash Med Squire)
Party ID Mark	AH/CF/02	Ref. Standard / Method	IS: 12508
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-S-04

Environmental conditions (s) : Temperature : (25 ± 10) °C. Relative Humidity : (50 ± 15) % RH

### Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	PCS/DTMC/01 / 109-0595Y	Transcal	TSC/22-23/18608-1	14.03.2024

### Calibration Results:

UUC Value (RPM)	Avg. Std. Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
500	500.6	-0.6	1.60%
1000	1000.8	-0.8	1.60%
1500	1501.3	-1.3	1.60%
2000	2001.9	-1.9	1.60%
2800	2802.3	-2.3	1.60%

Remark :-

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD  
DY.(QM/TM)

\*\*\*\*\* END OF CERTIFICATE \*\*\*\*\*



Note : The Certificate has issued subject to conditions stated back side overleaf