



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



CC-2339

Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No.	: CC233923000003893F	Calibration Done on	: 26.12.2023	S.R.F.No.	: 0002/26
Issue Date	: 28.12.2023	Calibration Valid Upto	: 25.12.2024	S.R.F Date	: 26.12.2023
UUC Receipt Date	: N/A	Certificate No.	: PCS/3893	Page	: 01 of 01
Calibrated for M/S. AAKASH HOSPITAL 90/43, Malviya Nagar, New Delhi-110017					

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Indicator With Sensor of Refrigerator	Size/Range	(-20 to 8) °C
Least Count	0.1 °C	Condition of Item	Good
Manufactured By	Whirpool	Calibration Performed at	Site
Manufacturer M/Sr. No.	N/A	Location	Lab
Party ID Mark	AH/RF/01	Ref. Standard / Method	DKD-R-5-7 / Comparison
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/02, Section No.CPTH-10 (S)

Environmental condition (s) : Temperature : (25 ± 10) °C Relative Humidity : (50 ± 20)% RH

Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
PT100X1, SIMPLEX, 4 - Wire RTD With Indicator	PCS/RTD/01 / 20C107440	Transcal	TSC/22-23/18608-3	14.03.2024

Calibration Results:

UUC Value	Avg. Std. Value	Error	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
°C	°C	°C	°C
-20.0	-19.524	-0.476	1.80
2.0	2.145	-0.145	1.80
4.0	4.251	-0.251	1.80
8.0	8.352	-0.352	1.80

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY.(QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No.	: CC233923000003824F	Calibration Done on	: 25.12.2023	S.R.F.No.	: 0001/25
Issue Date	: 28.12.2023	Calibration Valid Upto	: 24.12.2024	S.R.F Date	: 25.12.2023
UUC Receipt Date	: 25.12.2023	Certificate No.	: PCS/3824	Page	: 01 of 01
Calibrated for M/S. AAKASH HOSPITAL 90/43, Malviya Nagar, New Delhi-110017					

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Thermometer With Probe	Size/Range	(-50 to 300) °C
Least Count	0.1 °C	Condition of Item	Good
Manufactured By	R-Tek	Calibration Performed at	Lab
Manufacturer M/Sr.No.	RT-500	Location	Lab
Party ID Mark	AH/Temp./02	Ref. Standard / Method	ITS-90/Comparison
		Acceptance Criteria	N/A
Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/02, Section No.CPTH-07			

Environmental condition (s) : Temperature : (25 ± 2) °C. Relative Humidity : (50 ± 10) % RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
PT100X1, SIMPLEX, 4 - Wire RTD With Indicator	PCS/RTDI/01 / 20C107440	Transcal	TSC/22-23/18608-3	14.03.2024

Calibration Results: Set Points - (-10, 50, 150, 250) °C

UUC Value	Avg. Std. Value	Error	Related Uncer. on C.F.(k=2) at 95% Confidence Level
°C	°C	°C	±°C
-10.0	-10.113	0.113	0.64
50.0	50.126	-0.126	0.64
150.0	150.242	-0.242	0.64
250.0	250.369	-0.369	0.64

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233923000003825F	Calibration Done on : 25.12.2023	S.R.F.No. : 0001/25
Issue Date : 28.12.2023	Calibration Valid Upto : 24.12.2024	S.R.F Date : 25.12.2023
UUC Receipt Date : 25.12.2023	Certificate No. : PCS/3825	Page : 01 of 01
Calibrated for M/S. AAKASH HOSPITAL 90/43, Malviya Nagar, New Delhi-110017		

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Thermometer With Probe	Size/Range	(-50 to 300) °C
Least Count	0.1 °C	Condition of Item	Good
Manufactured By	R-Tek	Calibration Performed at	Lab
Manufacturer M/Sr.No	RT-500	Location	Lab
Party ID Mark	AH/Temp./01	Ref. Standard / Method	ITS-90/Comparison
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/02, Section No.CPTH-07

Environmental condition (s) : Temperature : (25 ± 2) °C. Relative Humidity : (50 ± 10) % RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
PT100X1, SIMPLEX, 4 - Wire RTD With Indicator	PCS/RTDI/01 / 20C107440	Transcal	TSC/22-23/18608-3	14.03.2024

Calibration Results: Set Points - (-10, 50, 150, 250) °C

UUC Value	Avg. Std. Value	Error	Related Uncer. on C.F.(k=2) at 95% Confidence Level
°C	°C	°C	±°C
-10.0	-10.117	0.117	0.64
50.0	50.129	-0.129	0.64
150.0	150.247	-0.247	0.64
250.0	250.358	-0.358	0.64

Remark :--

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf