



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2024

Lab Name	DR D.P RASTOGI CENTRAL RESEARCH INSTITUTE FOR HOMOEOPATHY	Lab No	18802
Constituent Group	Chemistry I	Date of Result Entered :	20/06/2024
PT item	Lyophilized human serum based	Date of Report Published :	27/06/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	D.BILIRUBIN	Diazonium salt ( Colorimetric ) / Jendrassik	Roche c111, c311, Integra 400, c501, c502, c503, c702	131	1.29	10.64	0.14	1.3 mg/dL	0.07	0.02
2	LDL	Direct method / Enzymatic colorimetric	Roche c111, c311, Integra 400, c501, c502, c503, c702	94	58.80	4.90	2.88	62 mg/dL	1.11	0.59

u\* - Method of Uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

**LAB ADDRESS :**

DR D.P RASTOGI CENTRAL RESEARCH INSTITUTE FOR HOMOEOPATHY  
A-1/1, SECTOR - 24  
NOIDA  
UTTAR PRADESH 201301

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*