



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2024



PC-1024

Lab Name **DR D.P RASTOGI CENTRAL RESEARCH INSTITUTE FOR HOMOEOPATHY** Lab No **18802**
 Constituent Group **HbA1c** Date of Result Entered : **18/06/2024**
 PT item **Lyophilized human whole blood based** Date of Report Published : **27/06/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche c111, c311, Integra 400, c501, c502, c503, c702	133	5.29	11.66	0.62	5.1 %	-0.31	0.11

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

DR D.P RASTOGI CENTRAL RESEARCH INSTITUTE FOR HOMOEOPATHY
A-1/1, SECTOR - 24
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****