



# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmmedicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION


Customer Name	CURA HOSPITAL	Certificate No.	GCMT/24/CH-02
		Date of Calibration	22/05/2024
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT, UTTARAHALLI HOBLI, BENGALURU, KARNATAKA- 560062	Valid Upto	21/05/2025
		Date of Issue	24/05/2024

Equipment Name	Make	Model	Serial Number	Department
ELECTROLYTE ANALYZER	SENSACORE	ST200	STAQW-5265	LABORATORY


Electrical Safety Test		
Description	Measured Value	Status
Main Voltage - Live to Neutral	229.6	Pass
Main Voltage - Live to Earth	229.8	Pass
Main Volt Neutral to Earth	0.4	Pass
Equipment Current	0.1	Pass
Earth Leakage Current	0.5	Pass

**Note:**

- This result of calibration refers to only to the particular items submitted for calibration.
- The above result is valid at the time of and under the stated conditions measurement.
- Decision rule – Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy.  
& "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
  
Dhanush  
Calibration Engineer



Authorized By:  
  
Harish  
Technical Director





COMMITTED TO THE CUSTOMER SINCE - 1996

# VAIDYANATHESHWARA INSTRUMENTS PRIVATE LIMITED

## CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

### 1 Name and Address of Customer

M/s.: Cura Hospitals,  
No.18, Kanakapura Main Road, Doddakallasandra  
Industrial Area, Uttarahalli Hobli, Bangalore - 560 062.

### 2 Customer Reference

- 2.1 ULR No : CC247324300003879F
- 2.2 Format No : VI-FRM-TH-002
- 2.3 SRF No : 2274
- 2.4 DC.No / Date : Ref.Letter / 12-06-2024
- 2.5 Receipt Date : 13-06-2024
- 2.6 Certificate No. : VI/24-25/2274-02
- 2.7 Issue Date : 15-06-2024

### 3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Digital Thermo Hygrometer (In / Out)
- 3.2 Make / Model : BEETECH / 288CTH-2
- 3.3 ID.No. : DTHM-01
- 3.4 DUC Condition : Satisfactory
- 3.5 Calibration Procedure No. / Ref. Std. : SOP-38-02 / Based On Comparison Method
- 3.6 No.of Pages : 2
- 3.7 Calibration Date : 15-06-2024
- 3.8 Calibration Due : 14-06-2025
- 3.9 Calibration done at : VI Thermal Lab
- 3.10 Discipline : Thermal

### 4 Environmental Condition:

Temperature: 25.3 - 25.7°C Humidity: 49 - 51 %RH

### 5 Standards Used for calibration:

Sl. No	Nomenclature	Make & Model	SI. No	Certificate No.	Cal Agency / Validity
1	Hygropalm With Sensor	Rotronic & HP23	60885044(Indicator) & 20146871(Sensor)	ETL/1047/23/C/F/275	FCRI, Palakkad / 15-09-2024
2	Platinum Resistance Thermometer With Indicator	Fluke/5609 (Sensor) & Isotech/Milli K (Indicator)	03520(Sensor) & 42202-11(Indicator)	FL/C/TH/20072023-C050	Fare Labs, Gurugram. / 02-08-2024
3	Humidity+Temperature Chamber And Dry Block Calibrator is Used As a Source				

### 6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133


Calibrated By

Poornima V  
(Calibration Engineer)

Checked By

Umesh D  
(Lab In-Charge)

Authorized By

  
Gangadhar C.K





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmmedicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	COVA HOSPITAL	Certificate No.	GCMT/24/CH-09
		Date of Calibration	22/05/2024
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOJI, HITTANAHALI, HOBLI, BENGALURU, KARNATAKA-560062	Valid Upto	21/05/2025
		Date of Issue	24/05/2024

Equipment Name	Make	Model	Serial Number	Department
BIOCHEMISTRY ANALYZER	GEN WORKS	AUTOCHEM EXP	NA	LABORATORY

Electrical Safety Test		
Description	Measured Value	Status
Main Voltage - Live to Neutral	229.8	Pass
Main Voltage - Live to Earth	229.4	Pass
Main Volt Neutral to Earth	0.3	Pass
Equipment Current	0.4	Pass
Earth Leakage Current	0.3	Pass

Note:

- > This result of calibration refers to only to the particular items submitted for calibration.
- > The above result is valid at the time of and under the stated conditions measurement.
- > Decision rule - Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy. & "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
*Dhanush*  
Dhanush  
Calibration Engineer



Authorized By:  
*Hansh*  
Hansh  
Technical Director

Website : [www.gcmmedicaltechnologies.com](http://www.gcmmedicaltechnologies.com)

GENWORKS HEALTH PVT LTD

Gamma Block, 5th Floor, Sigma Tech Park, Whitefield Main Road, Varthur Hobli, Bangalore 560066, India






# Vaidyanatheshwara Instruments Pvt Ltd.

E-Invoice

Calibration Center - One stop solution for all your Calibration Needs  
**ACCREDITED CALIBRATION LAB with vide certificate; CC-2473**

Telephone :080-23377266 Mob :8123340256 / 8884667063  
E-mail :accounts@viplgroup.com Web : www.viplgroup.com

**Unit 1, Registered Corporate Office & Works**  
No 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhinagar  
JB Kaval, Nandhini Layout Post, Bengaluru-560096



GSTIN : AAJCV8148G1ZR

(ORIGINAL FOR RECEIPT)

## Service Invoice/ Tax Invoice

<b>Consignee (Ship to)</b> <b>Cura Hospital</b> (Unit Of Ambara)No ,18 , kanakapura main road Doddakallsandra industrialarea.uttarahalli hobli Bengalore-560062 GSTIN/UIN : 29ABCFA3084B1ZS State Name : Karnataka		<b>Invoice No.</b> <b>VI/271/24-25</b>	<b>E Way Bill Number</b> 	<b>Invoice Dated</b> <b>15-Jun-24</b>
		<b>PO Number</b> <b>VI/24-25/QUOTP/May -052P</b>	<b>PO Dated</b> <b>25-May-24</b>	<b>SRP No</b> <b>VI/24-25/1016</b>
		<b>D.C. No</b> 	<b>D.C Dated</b> 	<b>Contact Person</b> <b>Mr. Purshottam</b>
<b>Buyer (Bill to)</b> <b>Cura Hospital</b> (Unit Of Ambara)No ,18 , kanakapura main road Doddakallsandra industrialarea.uttarahalli hobli Bengalore-560062 GSTIN/UIN : 29ABCFA3084B1ZS State Name : Karnataka		<b>Customer D.C No</b> 	<b>Customer D.C Date</b> 	<b>Contact Number</b> <b>9902234662</b>
<b>IRN</b> : 12711dd9d58587625dca3a853af191ddca0da16fb937-3479a8eebdea0				
<b>Ack No</b> : 112420691076465 <b>Ack Date</b> : 15-Jun-24				

SI No.	Description	HSN/SAC	Qty	Remarks	Service Charges	Calibration Charges	Amount
1	Refrigerator (Simple Calibration)	998346	2 nos	Calibration		1,000.00	2,000.00
2	Master Transportation Charges for Onsite	998346	1 nos	Calibration		950.00	950.00
	Sub Total						2,950.00
	Discount					(-)7.50 %	(-)221.25
	Amount After Discount						2,728.75
	OUTPUT CGST @ 9%						245.59
	OUTPUT SGST @ 9%						245.59
	Round Off						0.00
	<b>Total</b>		<b>3 nos</b>				<b>3,220.00</b>

Amount (in words): Three Thousand Two Hundred Twenty Only

Invoice No : AAJCV8148G  
 S.No : YAM-KR-03-0352219  
 Terms & Conditions:  
 1. Payment: Cheque/DD (Charges to your a/c) in favour of M/s. VAIDYANATHESHWARA INSTRUMENTS PVT LTD Payable at Bangalore.  
 2. No Claim or rejection will be entertained 15 days receipt of this bill.  
 3. Credit Period 30 Days from the Delivery.  
 4. Subject to Bangalore Jurisdiction.  
 5. The Liability of the laboratory is limited to the invoice amount only.

**Company's Bank Details:**  
 Bank Name : VAIDYANATHESHWARA INSTRUMENTS PRIVATE LIMITED  
 Bank : HDFC Bank (India)  
 Current A/c No. : 999999865867  
 IFS Code : HDFC0001232  
 Branch : Peenya Industrial Area



Scan to pay

For VAIDYANATHESHWARA INSTRUMENTS PRIVATE LIMITED 24-25







# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	CURA HOSPITAL	Certificate No.	GCMT/24/CH-03
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT, UTTARAHALLI HOBLI, BENGALURU, KARNATAKA- 560062	Date of Calibration	22/05/2024
		Valid Upto	21/05/2025
		Date of Issue	24/05/2024


Equipment Name	Make	Model	Serial Number	Department
CELL COUNTER	AVECON	HMT-60P	EC302207JVN04P	LABORATORY

### Electrical Safety Test


Description	Measured Value	Status
Main Voltage - Live to Neutral	227.3	Pass
Main Voltage - Live to Earth	228.2	Pass
Main Volt Neutral to Earth	0.2	Pass
Equipment Current	0.5	Pass
Earth Leakage Current	0.4	Pass

#### Note:

- This result of calibration refers to only to the particular items submitted for calibration.
- The above result is valid at the time of and under the stated conditions measurement.
- Decision rule – Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy. & "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
  
Daanush  
Calibration Engineer



Authorized By:  
  
Harish  
Technical Director





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmedicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	CURA HOSPITAL	Certificate No.	GCMT/24/CH/04
		Date of Calibration	22/05/2024
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT, UTTARAHALLI HOBLI, BENGALURU, KARNATAKA- 560062	Valid Upto	21/05/2025
		Date of Issue	24/05/2024


Equipment Name	Make	Model	Serial Number	Department
BIOCHEMISTRY SEMI-ANALYZER	AVECON	MAXLYZERSMART	ES100221	LABORATORY

### Electrical Safety Test

Description	Measured Value	Status
Main Voltage - Live to Neutral	230.3	Pass
Main Voltage - Live to Earth	228.4	Pass
Main Volt Neutral to Earth	0.5	Pass
Equipment Current	0.4	Pass
Earth Leakage Current	0.3	Pass

**Note:**

- This result of calibration refers to only to the particular items submitted for calibration.
- The above result is valid at the time of and under the stated conditions measurement.
- Decision rule – Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit/accuracy & "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
  
Dhanush  
Calibration Engineer



Authorized By:  
  
Harish  
Technical Director





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmmedicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJP17451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	CURA HOSPITAL	Certificate No.	GCMT/24/CH-07
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT,UTTARAHALLI HOBLI, BENGALURU, KARNATAKA-560062	Date of Calibration	24/05/2024
		Valid Upto	21/07/2025
		Date of Issue	24/05/2024


Equipment Name	Make	Model	Serial Number	Department
CENTRIFUGE	REMI	R-8C	2HIN-37129	LABORATORY

### Result


Sn	Set Values in RPM	Measured Values in RPM
1	500	496
2	1000	998
3	2000	1999
4	3000	
4	3500	

#### Note:

- > This result of calibration refers to only to the particular items submitted for calibration.
- > The above result is valid at the time of and under the stated conditions measurement.
- > Decision rule – Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy.  
& "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
  
Dhanush  
Calibration Engineer



Authorized By:  
  
Harish  
Technical Director





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagara, T. Dasarahalli,  
Bangalore-560057, E-mail : gcmicaltechnologies@gmail.com  
Phone : 9844903040, 9036689865, GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	CURA HOSPITAL	Certificate No.	GCMT/24/CH-05
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT, UTTARAHALLI HOBBLI, BENGALURU, KARNATAKA-560062	Date of Calibration	22/05/2024
		Valid Upto	21/05/2025
		Date of Issue	24/05/2024

Equipment Name	Make	Model	Serial Number	Department
MICROSCOPE	LABOMER	CXL PLUS	190314027	LABORATORY

## Electrical Safety Test

Description	Measured Value	Status
Main Voltage - Live to Neutral	229.3	Pass
Main Voltage - Live to Earth	229.4	Pass
Main Volt Neutral to Earth	0.3	Pass
Equipment Current	0.2	Pass
Earth Leakage Current	0.3	Pass

Note:

- > This result of calibration refers to only to the particular items submitted for calibration.
- > The above result is valid at the time of and under the stated conditions measurement.
- > Decision rule - Simple acceptance where "Pass" indicates measured value is within/equal to tolerance limit / accuracy. & "Fail" indicates measured value is out of tolerance limit / accuracy, without considering measurement uncertainty.

Calibrated By:  
  
Dhanush  
Calibration Engineer



Authorized By:  
  
Harish  
Technical Director





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, 1st Stage, Nagara, T. Dasarahalli,  
Bangalore-560057, E-Mail: [gcmedicaltechnologies@gmail.com](mailto:gcmedicaltechnologies@gmail.com)  
Phone : 9844903040, 9036000000 GSTIN : 29BCJPJ7451A1ZN

## CERTIFICATE OF CALIBRATION

Customer Name	CURA HOSPITAL	Certificate No.	GCMT/CH/CH-06
Address	NO 18, KANAKPURA ROAD, VENKATA LAYOUT, UTTARAHALLI HOBLI, BENGALURU, KARNATAKA-560062	Date of Calibration	22/05/2024
		Valid Upto	21/05/2025
		Date of Issue	24/05/2024

Equipment Name	Make	Model	Serial Number	Department
SEMI COAGULATION ANALYZER	TULIP DIAGNOSTIS	HEMOSTAR XF10	1232255445	LABORATORY

### Electrical Safety Test

Description	Measured Value	Status
Main Voltage - Live to Neutral	229.3	Pass
Main Voltage - Live to Earth	229.1	Pass
Main Volt Neutral to Earth	0.3	Pass
Equipment Current	0.2	Pass
Earth Leakage Current	0.3	Pass

Note:

- > This result of calibration refers to only to the particular items submitted for calibration.
- > The above result is valid at the time of and under the stated conditions measurement.
- > Decision rule - Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy & "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:  
*Dhanush*  
Dhanush  
Calibration Engineer



Authorized By:  
*Harish*  
Harish  
Technical Director





# GC MEDICAL TECHNOLOGIES

No.60, House #1533, 12th Cross, Prashanth Nagar, T. Dencorohalli, Bangalore-560057, E-mail : gcmmedicaltechnologies@gmail.com  
Website : www.gcmmedicaltechnologies.com  
Phone : 9844903040, 9036689865, GSTIN : 29BAGD00012N

## CERTIFICATE OF CALIBRATION

Customer Name	CURA MULTISPECIALITY HOSPITAL	Page No.	1 of 2
Address	Kanakapura Road, Doddakallasandra, Uttarahalli Hobli, Bengaluru, Karnataka-560062	Certificate No.	GCMT/23/016/164
		Date of Calibration	26/02/2023
		Valid Upto	26/02/2024
		Date of issue	05/03/2023

Equipment Details			
Equipment Name	Make	Model	Serial Number
Blood Roller Mixer	Precimix		NA
			Asset Number
			CHK/ME/LAB/164
			Department
			Laboratory

### Calibration Standards

Test Equipment	Manufacturer	Model	Serial Number	Cal Date	Due Date
Electrical Safety Analyzer	Fluke	ESA 612	5928535	21/09/2022	21/09/2023

### Calibration Status

Pass
  Fail
  Report Attached

#### Note:

- > This result of calibration refers to only to the particular items submitted for calibration.
- > The above result is valid at the time of and under the stated conditions measurement.
- > Decision rule - Simple acceptance: where "Pass" indicates measured value is within/equal to tolerance limit / accuracy & "Fail" indicates measured value is out of tolerance limit/accuracy, without considering measurement uncertainty.

Calibrated By:

*Kumar AN*  
Kumar AN  
Calibration Engineer



Authorized By:

*Chandrasekar*  
Chandrasekar  
Technical Director