



CRISPR BIOTECH

1/106, M-BLOC, B.P. TOWNSHIP, PATULI, KOLKATA-700094

CONTACT NO- 9874162850

Email: crispr.biotech2021@gmail.com

CALIBRATION CERTIFICATE

Date: 08/07/2024

This is to certify that **Mindray BS-240** Biochemistry Analyzer (Sl. No. **YX - 36007460**) installed at **LABAID DIAGNOSTICS PVT.LTD.** is calibrated as per the recommended procedures.

The system parameters as mentioned below were verified and necessary calibrations done as applicable.

System Temperatures:

Temperature Sensor	Temperature(°C)	Operatingrange(°C)
Reaction Carousel Temperature	37.0	36.7 to 37.3
Cuvette Cleaning Fluid Temperature	44.1	39.0 to 49.0
Cuvette Wash Solution Temperature	44.6	39.0 to 49.0

Power Supply Units:

Power Supply	Measured(V&A)	Range(V&A)
+24Volts	24.03V	20Vto30V
+12Volts	11.98V	11Vto14V
AC Mains Input Voltage to the instrument	224V	220Vto 240V
Ground Voltage	1.2V	<5Volts



CRISPR BIOTECH

1/106, M-BLOCK, B.P. TOWNSHIP. PATULI. KOLKATA-700094
CONTACT NO- 9874162850
Email: crispr.biotech2021@gmail.com

CALIBRATION CERTIFICATE

Date:08.07.2024

This is to certify that the **MindrayBS-240** Biochemistry Analyzer (Sl.No.YX - 36007460) installed at **LABAID DIAGNOSTICS PVT.LTD.** has been calibrated as per the recommended procedures. The system parameters are found to be within the specified ranges.

The System has passed the performance tests. The system is ready for calibration of the chemistries with the respective calibrators. The validation of the chemistry results must be done with multi levels of commercially available controls.

Calibration validity:12 Months

Last Calibration Done on: 08/07/2024
Next Calibration Due on: 07/07/2025

For,
CRISPR BIOTECH

Tamal Howlader
Service Manager
Mob : 89670-89439





CRISPR BIOTECH

1/106, M-BLOC, B.P. TOWNSHIP, PATULI, KOLKATA-700094

CONTACT NO- 9874162850

Email: crispr.biotech2021@gmail.com

Photometer Check:

Wavelength(nm)	Absorbance	Remarks
340	8506	Well within range
405	8652	Well within range
450	8589	Well within range
510	7581	Well within range
546	7511	Well within range
578	7564	Well within range
630	7496	Well within range
670	7531	Well within range

The equipment is certified to be in good working condition and is ready for the intended use.

For,
CRISPR BIOTECH

Tamal Howlader
Tamal Howlader
Service Manager
Mob : 89670-89439

