

BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.

NABL
ACCREDITED
LABORATORY

CALIBRATION CERTIFICATE

BI:CC.No:0433/2024

Name & Address : M/s. ICTC,
PHC SHAHEED DWEEP.

Master Instrument Reference : Tachometer, LT Make, DT-2236 Model,
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at
PCCPL, Chennai - 600 008. Report No.:PC-12-22/ATL/5194-01,
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)


Test Instrument Reference : CENTRIFUGE, REMI Make, Model: C-854/8,
Range: 0 to 3500RPM, Sl. No.: LN10-8890 (BI:0433/2024),
Identification No.: ICTC/PHC SHA/CENT-01.

Ambient Temp.When tested : 30°C

TEST READINGS


Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	± 1 %	rpm
506	500	"	-6
1014	1000	"	-14
2021	2000	"	-21
3039	3000	"	-39

Remarks: The above instruments well within the specified accuracy

Calibrated by : 
Calibrated on : 30-01-2024
Next Calibrn. Due on : 29-01-2025



For BHARATH INSTRUMENTS


Signatory