



Metropolis Healthcare Ltd., EQAS

AAKASH HOSPITAL-LAB (1090)

Outlier And Analyte Summary Report

Result Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Instrument : Erba EM 200

| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------------------|---------------|--------------|----------------|
| X C-Reactive Protein Qualitative | mg/L | Normal | Abn |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|-------------------------|-------------------------|
| ⓐ HBsAg | IU/mL | Reactive | Reactive |
| ⓐ HCV | IU/mL | Non-reactive | Non-reactive |
| ⓐ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| ⓐ HIV Ab | S/CO | HIV-1/HIV-2 Ab Reactive | HIV-1/HIV-2 Ab Reactive |
| ⓐ Typhidot IgM | - | Negative | Negative |
| ⓐ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| ⓐ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| ⓐ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| ⓐ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |

Legend ⓐ : Acceptable

ⓐ : Acceptable


X : Unacceptable

: Not Evaluated

ⓐ : Delayed Result Entry

| | |
|---------------------------------|---------|
| Total Parameters | 10 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 10 |
| Outlier Parameters (X) | 1 |
| EQAS Score Immunoassay Extended | 90.00 % |

Checked 
10.7.24

Note! Widal results satisfactory for this cycle. 
10.7.24


Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070

| Analyte | Instrument | Result Value | Standard Unit |
|--|-------------|--------------|---------------|
| X C-Reactive Protein Qualitative | Erba EM 200 | Normal | mg/L |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry | | | |

Problem Classification: All Parameters acceptable.
except CRP. Widal results satisfactory

Corrective Action: CRP done by Immunoturbidity Method
Quantitative but results option is only as
Qualitative i.e. Positive & Negative

Reviewed by: 

Dated: 10/07/24