

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification:


All Parameter of CBC & DLC is

acceptable. No outlier. TLC was unacceptable in last cycle. But acceptable this time

Corrective Action:

No C.A. required.


Reviewed by:



Dated:

10-07-24

Note: \Rightarrow TLC results are acceptable in this cycle. In previous cycle results ~~are~~ was entered ~~incorrectly~~ as per wrong units. So that only results was unacceptable.


10.7.24