



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 3650

Distribution No.: 161-J

Month/Year: October/2023

Instrument ID: Erba Elite 580

Model Name.: Erba Elite 580

Serial No.: K11051912060

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue & status of the report: 10-01-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	6.58	6.42	13	12.77	0.038	0.26	0.16	0.1	0.010	0.54
RBC x10 ⁶ /µl	1	4.24	4.06	8.3	8.72	0.011	-1.53	0.18	0.05	0.003	2.92
Hb g/dl	1	11	10.8	21.8	22	0.024	-0.34	0.2	0.1	0.008	1.35
HCT%	1	33.8	32.1	65.9	69.8	0.156	-0.99	1.7	0.4	0.027	2.92
MCV-fl	1	79.7	79.1	158.8	160	0.272	-0.19	0.6	0.3	0.025	0.67
MCH-Pg	1	26.5	25.9	52.4	50.45	0.063	1.38	0.6	0.2	0.016	1.80
MCHC-g/dl	1	33.5	32.5	66	63.3	0.146	0.73	1	0.3	0.022	1.89
Plt. x10 ³ /µl	1	187	181	368	397	1.557	-0.75	6	7	0.440	-0.17
Retic %	2	15.5	13.5	29	31	0.570	-0.14	2	1	0.066	0.90

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=0 , Poly=48 L=4, E=1, Mono/Promono=1 , B1=1 P.M.=7, Mye=23, Meta=12, Other=0
RBC Morphology	3	Poly: 43 - 56, Myelo: 14 - 28, Meta: 8- 16, Promyelo: 2-6, Lympho: 2- 5, Blast: 1-3, Eosino: 1-2, Mono: 1-2, nRBC/, Baso: 0-5
Diagnosis	3	Chronic myeloid leukemia (Chronic phase)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 161--J	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	277	273	82.05	89.74	5.86	4.03	12.09	6.23
RBC x10⁶/µl	1	277	277	88.09	90.25	6.14	2.89	5.77	6.86
Hb g/dl	1	277	277	84.12	90.25	6.14	2.89	9.74	6.86
HCT%	1	277	273	88.64	88.64	5.49	4.03	5.87	7.33
MCV-fl	1	277	273	87.18	91.58	6.96	2.56	5.86	5.86
MCH-Pg	1	277	272	83.82	93.01	7.72	1.1	8.46	5.89
MCHC-g/dl	1	277	273	90.48	89.01	4.4	4.03	5.12	6.96
Plt. x10³/µl	1	277	273	91.21	92.31	5.49	3.66	3.3	4.03
ReticCount%	2	277	237	96.62	91.14	2.95	3.38	0.43	5.48
PS Assessment	3	277	232	Satisfactory :95.68%, Borderline Sat. :2.88%, Unsatisfactory :1.44%					

***Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between " 0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ($0.3*SDPA$). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value ($0.3*SDPA$).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----