



Metropolis Healthcare Ltd., EQAS
HEALTH SWAYAM HSCP LABORATORIES PVT LTD, CHAKAN (864)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 03

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Urea	Mindray BS-240	24.70	mg/dl	2.05

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry
* : Not considered for evaluation.

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Mindray BS-240

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
! Urea	mg/dl	24.70	20.13	2.05	1.36
@ Creatinine	mg/dl	0.78	0.70	0.18	-0.36
@ Bilirubin Total	mg/dl	0.19	0.19	0.02	-0.35
@ Total Protein	g/dL	6.06	5.58	0.82	0.26
@ Albumin	g/dL	3.42	3.36	0.18	0.14
@ Uric Acid	mg/dl	4.10	4.13	-0.11	-0.70
@ Calcium	mg/dl	6.70	7.25	-1.01	-0.20
@ Cholesterol	mg/dl	110.00	125.80	-1.28	-0.81
@ Triglycerides	mg/dl	109.90	111.99	-0.45	-0.21
@ HDL	mg/dl	31.10	31.99	-0.28	-0.20
@ ALT	U/L	13.00	14.59	-0.17	-0.08
@ AST	U/L	22.80	23.19	-0.44	-0.32
@ ALP	U/L	37.00	46.40	-1.00	-0.88
@ GGT	U/L	19.00	24.97	-1.91	0.16

Instrument : Sensa Core ST 200 plus

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	89.90	86.36	1.72	1.25
@ Sodium	mmol/L	99.80	112.00	-1.26	-0.12
@ Potassium	mmol/L	3.42	3.55	-0.29	-0.52

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	17
Not Evaluated Parameters	0
Evaluated Parameters	17
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070