



# ARK Diagnostics Bangalore Pvt. Ltd.

## CERTIFICATE of CALIBRATION

**Manufacturer : ARK DIAGNOSTICS BANGALORE PVT. LTD.,**

**Instrument : HPLC ANALYSER**

**Serial Number : CKK-A1C#123-006**

**Customer : SRI ADICHUNCHUNAGIRI HOSPITALS & RESEARCH CENTRE**

Ark Diagnostics Bangalore Pvt. Ltd., certifies that the above listed instrument was calibrated on date noted at, **BALAGANGADHARANATHA NAGAR, NAGAMANGALA TALUK, MANDYA DIST- 571448**, in accordance with the specifications set forth by the manufacturer using the certified control material as per the provided assay values and expected ranges from the manufacturer.

Calibration Date:-**12th DEC 2023**

Calibrated by **Mr.LITHIN**

Calibration Due:-**11th DEC 2024.**

**For ark Diagnostics Bangalore Pvt. Ltd.,**



\*Subjected to AMC



## INSTALLATION QUALIFICATION

**INSTRUMENT NAME :** CKKA1c  
**SERIAL NUMBER :** 123-006  
**CUSTOMER NAME :** Sri Adichunchunagiri Central Laboratory – Mandya.  
**ADDRESS :** Balagangadhara Nagar , Nagamangala Taluk , Manyda  
Pin-571448.

### INSTALLATION REQUIREMENTS FOR THE INSTRUMENT:

REQUIREMENTS	ACCEPTABLE RANGE	OBSERVED RANGE
Input Voltage	230+10AC	232+5V AC
Line Frequency	50/60 HZ	50HZ
Ambient Temperature	15-45 degree	21 degree

**FOR ARK DIAGNOSTICS BANGALORE PVT. LTD.**






## INSTALLATION REPORT

**CUSTOMER NAME** : Sri Adichunchunagiri Central Laboratory – Mandya.  
**ADDRESS** : Balagangadhara Nagar , Nagamangala Taluk , Manyda  
Pin-571448.  
**DEPARTMENT** : CLINICAL BIOCHEMISTRY  
**CONTACT PERSON** : MR.NANDHISH  
**INSTRUMENT NAME** : CKKA1c  
**SERIAL NUMBER** : 123-006  
**INSTALLATION DATE** : 16/JAN/2019  
**WARRANTY EXPIRY** : 1 year

THE HbA1c ANALYZER WAS INSTALLED ALONG WITH THE NECESSARY STANDARD ACCESSORIES. ALL BASIC REQUIREMENTS AND ADJUSTMENTS WERE CHECKED AND FOUND TO BE SATISFACTORY. THE PRELIMINARY STANDARDIZATION OF THE HbA1c ANALYZER AND THE TRAINING WERE PROVIDED TO THE COMPLETE SATISFACTION OF THE CUSTOMER. THE HbA1c ANALYZER IS FOUND TO BE WORKING SATISFACTORILY.

FOR ARK DIAGNOSTICS BANGALORE PVT. LTD.

MR. NANDHISH  
(SERVICE ENGINEER)





**OPERATIONAL QUALIFICATION OF THE INSTRUMENT:**

DESCRIPTION	STATUS	REMARKS
POWER SUPPLY	SELF CHECK	PASSED
ELECTRONICS	SELF CHECK	PASSED
REAGENT PACK	SELF CHECK	PASSED
HARDWARE	SELF CHECK	PASSED
FLOW PATH	SELF CHECK	PASSED

**For Ark Diagnostics Bangalore Pvt. Ltd.**

  
  
(AUTHORIZED SIGNATORY)