



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.69	2.73	-0.19	-0.17
@ Total Leucocytes (WBC) count	/c.mm	7250.00	7841.43	-0.46	-0.20
@ Haemoglobin	g/dL	7.80	8.06	-0.86	-1.00
@ MCV (Mean Corpuscular Volume)	fL	121.80	117.59	0.23	0.25
@ MCH (Mean Corpuscular Hb)	pg	29.20	29.58	-0.14	-0.35
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.90	24.80	-0.48	-0.40
@ Platelet Count	1000/ μ L	293.00	271.07	1.21	1.69
@ Hematocrit	%	32.80	31.87	-0.36	-0.40

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 05

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	40-70%	40-70%
@ Lymphocyte.	-	20-40%	20-40%
@ Eosinophil.	-	<6%	<6%
@ Monocyte.	-	1-10%	1-10%

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 05

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value										
@ Diagnosis (Slide)	-	P. vivax	P. vivax										
Legend @ : Acceptable	<table border="1"> <tr> <td>Total Parameters</td> <td>1</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Haematology- Peripheral smear diagnosis</td> <td>100.00 %</td> </tr> </table>			Total Parameters	1	Not Evaluated Parameters	0	Evaluated Parameters	1	Outlier Parameters (X)	0	EQAS Score Haematology- Peripheral smear diagnosis	100.00 %
Total Parameters				1									
Not Evaluated Parameters				0									
Evaluated Parameters				1									
Outlier Parameters (X)				0									
EQAS Score Haematology- Peripheral smear diagnosis	100.00 %												
@ : Acceptable													
X : Unacceptable													
# : Not Evaluated													
⊙ : Delayed Result Entry													



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Platelet Count	Horiba Yumizen H550	187.00	1000/ μ L	2.17
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.44	2.46	-0.15	-0.17
@ Total Leucocytes (WBC) count	/c.mm	2770.00	2687.88	0.07	-0.20
@ Haemoglobin	g/dL	6.60	6.75	-1.13	-1.00
@ MCV (Mean Corpuscular Volume)	fL	110.20	106.16	0.27	0.25
@ MCH (Mean Corpuscular Hb)	pg	26.90	27.26	-0.55	-0.35
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	24.40	25.63	-0.32	-0.40
! Platelet Count	1000/ μ L	187.00	158.49	2.17	1.69
@ Hematocrit	%	26.90	26.12	-0.44	-0.40

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070