



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal
Legend @ : Acceptable		Total Parameters	1
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	1
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

TATA 1 MG LABS MERAMANDALI(TATA 1 MG TECHNOLOGIES PVT LTD)
(1247)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X pH	Clinitex status +	5.5	-
X Specific Gravity	Clinitex status +	1.020	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Clinitex status +


Analyte	Standard Unit	Result Value	Accepted Value
X pH	-	5.5	5
X Specific Gravity	-	1.020	1.025
@ Glucose (urine)	-	3+	3+
@ Protein (Urine)	-	2+	2+
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	2+	2+
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
@ Nitrite (urine)	-	Present	Pre
@ Leucocyte (Urine)	-	Negative	Negative

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Positive

Legend @ : Acceptable
@ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	2
EQAS Score Clinical Path - Urine Routine Chemistry	81.82 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070